

Simultaneous transdermal extraction of glucose and lactate from human subjects by reverse iontophoresis

Tak S Ching¹
Patricia Connolly²

¹Asia University, Taiwan;

²Bioengineering Unit, University of Strathclyde, Glasgow, Scotland, UK

Abstract: This study investigated the possibility of simultaneously extracting glucose and lactate from human subjects, at the same skin location, using transdermal reverse iontophoresis. Transdermal monitoring using iontophoresis is made possible by the skin's permeability to small molecules and the nanoporous and microporous nature of the structure of skin. The study was intended to provide information which could be used to develop a full, biosensor-based, monitoring system for multiple parameters from transdermal extraction. As a precursor to the human study, *in vitro* reverse iontophoresis experiments were performed in an artificial skin system to establish the optimum current waveforms to be applied during iontophoresis. In the human study, a bipolar DC current waveform (with reversal of the electrode current direction every 15 minutes) was applied to ten healthy volunteers via skin electrodes and utilized for simultaneous glucose and lactate transdermal extraction at an applied current density of 300 $\mu\text{A}/\text{cm}^2$. Glucose and lactate were successfully extracted through each subject's skin into the conducting gel that formed part of each iontophoresis electrode. The results suggest that it will be possible to noninvasively and simultaneously monitor glucose and lactate levels in patients using this approach and this could have future applications in diagnostic monitoring for a variety of medical conditions.

Keywords: transdermal, iontophoresis, glucose, lactate, diagnostic monitoring

Introduction

Epidemiological studies have shown that the prevalence of diabetes is steadily increasing and is a widespread problem in modern society (King et al 1993; Amos et al 1997; Wild et al 2004). The number of cases of diabetes worldwide in 2000 among adults ≥ 20 years of age was estimated to be 171 million (Wild et al 2004) and in 2003 this was estimated to be 194 million (IDF 2003). By 2025, the number of people with diabetes is expected to exceed 333 million.

Diabetes increases the risk of ill health and shortens life. It is the leading cause of blindness and visual impairment (Evans 1995), amputation (Davis et al 2004), cardiovascular disease (Laing et al 1999a, 1999b), stroke (Bell 1994; Stephenson et al 1995) and kidney failure (Wang et al 1996; Brancati et al 1997).

Tight monitoring and control of blood glucose levels can reduce the prevalence of complications in both Type 1 and Type 2 diabetes (UK Prospective Diabetes Study Group, 1998a, 1998b, 1998c, 1998d). To date, most Type 1 and Type 2 diabetic patients measure their own blood glucose several times a day by obtaining finger-prick capillary samples and applying the blood to a reagent strip for analysis in a portable meter (Pickup 2003). However, this method is painful, cumbersome, aesthetically unpleasant and inconvenient and blood glucose monitoring cannot continue during sleep. This provides considerable impetus for the development of noninvasive methods for continuous blood glucose monitoring.

Correspondence: Patricia Connolly
Bioengineering Unit, University of
Strathclyde, Glasgow G4 0NW, Scotland,
UK
Tel +44 141 548 3034
Email patricia.connolly@strath.ac.uk

Skin provides a unique gateway for noninvasive, transdermal monitoring. Recently, reverse iontophoresis techniques have been used for patient monitoring (Rao et al 1995; Tamada et al 1999; Pitzer et al 2001; Tierney et al 2001; Potts et al 2002; Degim et al 2003) and noninvasive diagnosis (Numajiri et al 1993; Mize et al 1997; Merino et al 1999), most notably the GlucoWatch® biographer which proposed a solution for diabetics which avoided the frequent need for blood sampling (Tamada et al 1999). Reverse iontophoresis refers to the passage of a low level of current through the skin, normally via a pair of skin electrodes containing a conducting gel, to promote the extraction of both charged and neutral molecules. The major mechanisms for transport of ions and molecules are; the electromigration of charged species to the electrode of opposite polarity, electroosmosis of neutral molecules to the cathode or anode, or a combination of these two processes. The most successful reverse iontophoresis application for patient monitoring to date, as mentioned, has been in noninvasive and continuous blood glucose detection utilising the GlucoWatch®. The GlucoWatch® biographer utilized the techniques of reverse iontophoresis and biosensing to noninvasively and continuously monitor blood glucose. However, the GlucoWatch® had some problems, for example it was reported that it does not reliably detect hypoglycemia in children (The Diabetes Research in Children Network 2004). In addition there have been no further developments of commercial systems to monitor any other blood analytes transdermally. Consequently there is room for further study, understanding and development of transdermal reverse iontophoresis for noninvasive patient monitoring.

The present study investigated the possibility that other blood parameters could be monitored simultaneously with glucose through the skin utilizing a single device on a patient. Lactate was selected as a parameter of interest for a variety of clinical conditions. Glucose and lactate levels are metabolically linked and therefore are indicative of the healthy functioning of an organism or tissue section when monitored simultaneously

Monitoring of lactate is known to be beneficial to the critical care patient as the trend in lactate levels is an indicator of patient viability (Abramson et al 1993). Blood lactate levels have been widely used for prognosis of critical illness and assessing effects of therapeutic maneuvers (Cairns et al 1997; Vincent 1998; Mooney et al 1999). Blood lactate is an extremely fast indicator of oxygen deprivation at the cellular level, increasing within seconds of a life threatening event and decreasing within seconds of appropriate therapy

(Bakker et al 1991). Blood lactate has also been reported to be superior to other variables (eg, oxygen-derived variables) as a predictor of outcome (Mooney et al 1999). In critical care units, high lactate levels in the general circulation have a strong prognostic value (Vincent 1996), with lactate levels over 5 mM (normally 0.6–2.4 mM) associated with a poor outcome. Therefore, it is possible to predict organ failure from serial lactate measurements (Gersh and Anderson 1993). Lactate monitoring has been suggested as valuable for many patients in emergency medicine: for surgical patients (Elia et al 1991; Hildingsson et al 1996; Carrier et al 1998; Habicht et al 1998), for trauma patients (Slomovitz 1998), for patients with head injury or cerebral ischemia (Valadka 1998; Menzel et al 1999), acute intestinal ischemia (Murray et al 1994), liver ischemia (Rasmussen et al 1994), transplanted organ surveillance eg, myocutaneous flaps (Rojdmark et al 1998), intrapartum for the fetus (Westgren et al 1999), and patients with septic shock (Boer et al 1994).

Clinically, a patient's blood lactate level can be tracked in a minimally-invasive manner by microdialysis (Volpe et al 1994) or through finger-stick sampling. Although the microdialysis sampling method, when combined with a biosensor or other analysis technique, can provide continuous lactate monitoring it still has a chance of inducing infection on the insertion site and surgical intervention is required for this technique. The microdialysis tubing generally needs replacement after a 48 hour period. On the other hand, the finger-stick method cannot provide continuous lactate monitoring.

The ability to monitor both glucose and lactate simultaneously and noninvasively would have applications in many of the clinical conditions described above. In addition this measurement technique could be applied to monitor the viability of skin grafts post operatively and to assess the vascular condition of skin and limbs.

The use of reverse iontophoresis in patient monitoring does present some concerns. For example, use of constant direct current (DC) during transdermal reverse iontophoresis may localise and generate hydroxyl ions in the anodal skin region and hydrogen ions in the cathodal skin region. This process, which disrupts local pH values, leads to stinging and erythema (Howard et al 1995). The use of pulsed DC, (a constant DC delivered periodically), has been suggested to prevent charge build up on the skin (polarization) and minimize pH changes (Chien et al 1987; Chien et al 1990). Polarization effects turn the skin into what is effectively a charged capacitor and this reduces the flow of current and charged species during iontophoresis. When current is

switched off the skin naturally depolarizes and returns to its near initial electrical condition. The use of bipolar current, a constant DC source where direction of current flow through the skin electrodes is changed periodically, has also been suggested as a means to overcome skin polarisation and minimize pH changes (Howard et al 1995; Tomohira et al 1997). Tomohira and colleagues (1997) used an *in vivo* rat abdominal skin model to study the effect of electrode polarity switching (ie, a bipolar DC current profile with the polarity of the electrodes switched over at specific intervals) in the iontophoresis of insulin and calcitonin for transdermal delivery. They found that applied bipolar DC current profiles could enhance the absorption of the insulin and calcitonin. Hirvonen and colleagues (1995) used a diffusion cell with mouse skin to study the impact of different applied current profiles (waveforms) in regulating the permeation of two charged amino acids, lysine and glutamic acid. They found that bipolar current waveforms or constant DC current waveforms resulted in comparable transport rates which are both higher than that of pulsed DC current waveforms. Bipolar current waveforms were selected for investigation in this study.

The aims of this study were; (1) to establish the optimum applied current waveforms for the transdermal reverse iontophoresis of both glucose and lactate and (2) to study the simultaneous reverse iontophoretic extraction of glucose and lactate across human skin barriers in healthy human subjects.

The effects of applied current waveforms on the reverse iontophoresis of glucose and lactate were initially investigated using *in vitro* diffusion cells designed to mimic human skin conditions. These were utilized with an artificial membrane as a skin substitute, avoiding the sample to sample variation found in human skin whilst allowing the study of how the electrical waveform affects the transport of glucose and lactate (Connolly et al 2002). The optimum applied current waveform found from these experiments was then applied to healthy human volunteers, and glucose and lactate were extracted simultaneously from the subjects' skin using a specially-designed collection electrode.

Materials and methods

Constant current source

A fully programmable, portable constant current source, designed and developed by the authors for iontophoresis applications (Ching et al 2005), was used to deliver iontophoretic current for both *in vitro* and human reverse iontophoresis experiments. The device is capable of delivering constant

DC current, bipolar DC current waveforms and DC current pulses according to how it is programmed.

In vitro reverse iontophoresis experiments

A HEPES buffer solution was used for *in vitro* experiments. N-2-hydroxyethylpiperazine-N'-2-ethanesulphonic acid (HEPES), sodium chloride (NaCl), sodium hydroxide (NaOH), hydrochloric acid (HCl) and lactate were purchased from Sigma Chemical Co. (St. Louis, MO). Glucose was purchased from BDH Limited (Poole, England). Lactate reagents (LC 2389) and glucose reagents (GL 26233) were purchased from Randox Laboratories Limited (Antrim, UK). De-ionized water (resistivity $\geq 18\text{M}\Omega\text{cm}$) that had been purified by a Millipore System (Milli-Q UFplus; Bedford, MA) was used to prepare all solutions. Nanoporous membrane with a net negative charge at pH 7 was used as the artificial skin model (Spectra/Por[®] cellulose ester dialysis membrane MWCO: 500, Spectrum Laboratories, Inc., Canada) in accordance with earlier studies of the iontophoresis model (Connolly et al 2002).

Silver-silver chloride (Ag/AgCl) electrodes were prepared by chloriding silver wire (1mm diameter, 25 mm length, 99.99% pure; Aldrich Chemical Company Inc., Milwaukee, WI) immersed in 0.1M HCl solution (Pt-cathode) for 90 minutes at an applied current of 314 μA .

All *in vitro* experiments were performed using model diffusion cells designed for iontophoresis experiments (Connolly et al 2002), in which both electrode chambers were located on the same surface side of a nanoporous membrane as shown in Figure 1. The electrode chambers were filled with 350 μl of 25 mM, pH 7.4, HEPES buffer containing 133 mM NaCl. The lower chamber of the diffusion cell contained an electrolyte solution comprising 133 mM NaCl, buffered to pH 7.4 with 25 mM HEPES, and either 5 mM glucose or 10 mM lactate. Each electrode chamber contained a Ag/AgCl electrode. The surface area of the nanoporous membrane exposed to the electrode in each chamber was 0.2 cm^2 and the electrode chambers were 11 mm apart. The device was operated at ambient temperature (21–23 $^{\circ}\text{C}$).

An iontophoretic current of 0.3 mA/cm^2 was used for the various applied current waveforms tested *in vitro* in the cells and each experimental current condition was sustained for 60 minutes with current supplied through the Ag/AgCl wire electrodes.

The applied current waveforms tested in this study were: (i) constant DC current for the entire duration of current flow, (ii) DC current with the polarity of the electrode pair




Figure 1 Schematic illustration of the model diffusion cell for *in vitro* reverse iontophoresis experiments. For the control experiment, the experimental setting was identical except no current flowed through the Ag/AgCl electrodes.

reversed every 5 minutes during the 60 minutes of the current flow, (iii) DC current with the polarity of the electrode pair reversed every 10 minutes during the 60 minutes of the current flow, (iv) DC current with the polarity of the electrode pair reversed every 15 minutes during the 60 minutes of the current flow.

Each applied current waveform was tested in the model iontophoresis cell and the entire contents of the electrode chambers were removed at the end of each experimental procedure and analyzed to quantify the amount of glucose or lactate extracted through the nanoporous membrane. For the control experiments, all the experimental arrangements and procedures were the same as those described for the reverse iontophoresis experiments except that no current was applied in the control experiments, allowing diffusion of lactate and glucose through the nanoporous membrane to be observed in the collection chambers.

A colorimetric clinical chemistry assay method, was adapted for use with a microplate reader (Multiskan Ascent[®], Labsystems Oy, Finland), and was used to quantify the amount of glucose (determined by Glucose reagents, GL 26233, Randox Laboratories Limited, UK) or lactate (determined by Lactate reagents, LC 2389, Randox Laboratories Limited, UK) extracted through the nanoporous membrane. For the test, 150 μ l of each extracted sample was mixed with 150 μ l of the corresponding Glucose/Lactate reagents and dispensed as a sample into a 96-well microplate (Multiskan Ascent, Labsystems Oy, Finland). When loaded with samples the microplate was then immediately placed on the microplate reader. Each plate was incubated for 90 minutes at 37 °C in

the microplate reader and absorbance read at either 500 nm (for glucose) or 550 nm (for lactate). Excellent linear relationships ($r^2 > 0.99$ for both cases) between lactate/glucose standard solutions and their relative absorbance were found in the calibration curves for these assays, allowing the lactate/glucose concentrations from the model cells to be calculated by linear regression.

Human reverse iontophoresis experiments

Sodium phosphate monobasic (USP grade) and NaOH were purchased from Sigma-Aldrich. Methylcellulose (MC), Methocel A4M Prem, from the Dow Chemical Company was used to form the electrode contact gels. 0.1 M phosphate buffer solution (PBS) was prepared by dissolving 1.1998 g of sodium phosphate monobasic in 100 ml de-ionized water and adjusting the pH to 7.4 using a 40% w/v solution of NaOH. A 4% w/v MC gel was prepared by mixing 4g of MC with 100 ml of 0.1M PBS. De-ionized water (resistivity ≥ 18 M Ω cm) that had been purified by a Millipore System (Milli-Q UFplus; Bedford, MA) was used to prepare all solutions.

Circular (11.3 mm diameter) screen-printed Ag/AgCl electrodes (SPE) were fabricated by printing ink through a screen pattern to a thin polyvinyl chloride sheet. Silver paste and Silver-silver chloride (Ag/AgCl) paste for printing the electrodes was purchased from Advanced Conductive Materials (Atascadero, CA). Each SPE was coated with 90 μ l of the MC (4%) gel described above within a silicone ring. The MC (4%) gel acts as a conductive media for the transmission of iontophoretic current from the SPE to the human skin and also acts as

a collector of the extracted lactate and glucose. The prepared electrodes were used in pairs on the human subjects. A prepared electrode ready for human use is shown in Figure 2.

Ten healthy human volunteers (8 men and 2 women), aged between 22 and 35 years (mean = 26.8, SD = 3.7), with no history of dermatological disease participated in this study. The subjects were required to maintain the forearm sites under investigation free from application of any cosmetic topical formulations for at least 5 days before the study. The subjects were not fasted before the study and the measurements were taken during the working day. The study was approved by the University of Strathclyde Ethics Committee and informed consent was obtained from each volunteer.

The areas of skin of each subject's inner forearm where SPEs were to be located were prepared by briskly rubbing the areas for 6–8 seconds with alcohol soaked pads to remove dry skin, oils and other contaminants. The skin was then allowed to dry thoroughly. Then, four SPEs were positioned on the subject's inner forearms and fixed in position with surgical tapes. Each pair of SPEs was set with about 23 mm between the electrode centres and the two pairs of SPEs were kept about 50 mm apart. An iontophoretic applied current waveform with a current density of 0.3 mA/cm², and polarity reversal of the electrodes every 15 minutes during current flow, was applied on one pair of the SPEs on each volunteer's skin for a period of 60 minutes. The second pair of SPEs was used as a simultaneous control with no current passing through it. At the end of the experiment, the gel sampled from each of the paired SPEs (2 controls and 2 current sources) was mixed thoroughly as it was sampled and 20 µl samples

of the MC gel were carefully pipetted from each SPE and stored separately in microcentrifuge tubes at 4 °C for later quantification of lactate and glucose. The subject's blood glucose and lactate level was also measured before and after experiments by a portable glucose meter (FreeStyle Blood Glucose Monitoring System, TheraSense Ltd., UK) and a portable lactate meter (Accutrend® Lactate, Roche Diagnostics GmbH, Germany), respectively.

For the analysis of the samples 180 µl of 0.1M PBS was pipetted to each microcentrifuge tube containing the extracted 20 µl MC gel. The solutions were then mixed well. These solutions were then subjected to the colorimetric assays described before to determine the concentration of lactate and glucose in each gel sample.

Results

In vitro studies

In vitro diffusion studies were performed using the nanoporous membrane as the skin-like barrier in the iontophoresis cell with and without application of iontophoretic current. It was shown that passage of a current facilitates movement of lactate and glucose across the nanoporous membrane when compared to ordinary diffusion. The applied electrical current waveforms for reverse iontophoresis of lactate and glucose were (i) DC only and (ii) DC with an electrode polarity reversal every 5, 10 or 15 minutes, respectively. The results for glucose and lactate *in vitro* extraction are shown in Figure 3.

In the *in vitro* lactate diffusion studies, it was found that there is a significant difference between the four applied




Figure 2 The screen-printed electrodes. (a) The front view, with methylcellulose gel (4%) pipetted to this side. (b) The back view, with circular window for crocodile clip connection.




Figure 3 *In vitro* reverse iontophoresis extraction of (a) lactate and (b) glucose (mean \pm SD; $n \geq 12$ for each bar), as a function of applied current waveforms. The iontophoretic current and current application time are 0.3 mA/cm² and 60 minutes, respectively. The electrolyte in the electrode chambers of the diffusion cell was 25 mM, pH 7.4, HEPES buffer containing 133 mM NaCl. The lower chamber of the diffusion cell were filled with an electrolyte solution comprising 133 mM NaCl, buffered to pH 7.4 with 25 mM HEPES, and either 10 mM lactate or 5 mM glucose.

current waveforms (one-way ANOVA: $p < 0.001$). Lactate extraction at the anode during continuous current passage (without a polarity reversal) was, as might be expected for a negatively charged ion, found to be significantly higher than that at all other applied current waveforms (LSD post-hoc multiple comparisons: $p < 0.001$ in all cases), except the 15 minute electrode polarity reversal waveform. Lactate

extraction using the 15 minute electrode polarity reversal condition was found to be significantly higher (LSD post-hoc multiple comparisons: $p < 0.001$ in both cases) than that of protocols utilizing 5 and 10 minutes electrode polarity reversal conditions. It was also found that there was a significant higher lactate extraction for 5, 10 and 15 minutes electrode polarity reversal protocols than that in the control sample

which utilized diffusion-only extraction (LSD post-hoc multiple comparisons: $p < 0.05$ in all cases).

In the *in vitro* glucose diffusion studies, a significant difference was found between the four applied current waveforms (one-way ANOVA: $p < 0.001$). It was found that glucose extraction at 15 minute electrode polarity reversal was significantly higher than that at all other applied current waveforms (LSD post-hoc multiple comparisons: $p < 0.001$ in all cases). A significantly higher glucose extraction was also found in 5, 10 and 15 minutes electrode polarity reversal than the control (LSD post-hoc multiple comparisons: $p < 0.001$ in all cases). It can be seen in Figure 3 that anodal extraction of glucose was greater than cathodal extraction but it was not a significant difference.

Transdermal extraction of glucose and lactate from human subjects

During the human studies, lactate and glucose were successfully extracted simultaneously from the subjects' skin using reverse iontophoresis.

Only one subject (Subject D) reported any discomfort, mainly the experience of a very weak tingling sensation as the current was brought to 0.3 mA at the start of the experiment and when the electrode polarity reversed. At time zero (the start of current flow), she experienced a tingling sensation which lasted no longer than two seconds. At time = 15 minutes of current passage (ie, the moment of electrode polarity reversal), she again experienced a tingling sensation which also lasted no longer than two seconds. After that, she did not experience any other further discomfort or sensations. Four subjects had very mild erythema at the reverse iontophoresis

site after the experiment and this lasted for 15–30 minutes after termination of current flow.

The results of analysis of the human studies are summarized in Tables 1 and 2 and Figures 4, 5 and 6. Transdermal diffusion of both lactate and glucose occurred for each subject, as observed in the control electrodes that were not subjected to iontophoretic current. However, reverse iontophoresis was found to significantly promote more glucose extraction (around 4 times greater) and lactate extraction (around 2.5 times greater) than diffusion alone (independent t-test: $p < 0.001$ in both cases) as shown in Figure 4.

The blood glucose levels of the ten healthy volunteers at the start of the experiment and the extracted glucose levels (averaged glucose found in both gel electrodes) after reverse iontophoresis were compared (see Figure 5) and poor correlation was found between extracted and blood glucose across the subject group ($r^2 = 0.44$). It was found that there was no significant difference in the volunteers' blood glucose levels before and after experiment (paired t-test: $p = 0.227$). The poor correlation across the group when comparing internal versus extracted glucose is due to skin permeability differences between human subjects and an individual calibration of the device for each subject would be required for practical, time dependent monitoring.

Similarly, the blood lactate levels of the ten healthy volunteers at the start of the experiment and the extracted lactate levels (averaged lactate extraction across both electrodes) were also compared (see Figure 6) and no correlation was found ($r^2 = 0.19$). It was found that there was no significant difference on the volunteers' blood lactate level before and after experiment (paired t-test: $p = 0.116$). Again, the lack

Table 1 Measured blood lactate level and transdermally extracted lactate level for the 10 healthy volunteers

	Real blood lactate level (mmol/L)		Extracted lactate level (μM)							
	Before experiment	After experiment	Reverse iontophoresis group				Control group			
			Electrode 1	Electrode 2	Mean	SD	Electrode 1	Electrode 2	Mean	SD
Subject A	1.8	1.1	426.9	350.2	388.5	54.2	109.2	107.1	108.2	1.5
Subject B	1.6	1.3	374.9	360.0	367.5	10.5	123.5	115.0	119.2	6.0
Subject C	2.2	1.7	311.5	315.7	313.6	3.0	181.8	132.9	157.4	34.6
Subject D	1.0	1.2	361.7	238.1	299.9	87.4	182.7	116.6	149.6	46.8
Subject E	2.4	1.8	377.5	460.6	419.0	58.7	128.2	128.2	128.2	0.0
Subject F	1.4	1.9	347.9	369.2	358.5	15.1	128.7	122.3	125.5	4.5
Subject G	1.9	1.5	372.8	285.7	329.2	61.6	151.7	200.6	176.2	34.6
Subject H	1.5	1.6	415.1	344.7	379.9	49.8	180.5	140.0	160.2	28.6
Subject I	1.6	1.5	408.0	295.0	351.5	79.9	158.5	149.9	154.2	6.1
Subject J	1.5	1.2	365.7	312.6	339.2	37.5	108.9	104.6	106.8	3.0
Mean	1.7	1.5	376.2	333.2			145.4	131.7		
SD	0.4	0.3	34.0	59.8			29.5	28.1		

Table 2 Measured blood glucose level and transdermally extracted glucose level for the 10 healthy volunteers

	Real blood glucose level (mmol/L)		Extracted glucose level (μM)							
	Before experiment	After experiment	Reverse iontophoresis group				Control group			
			Electrode 1	Electrode 2	Mean	SD	Electrode 1	Electrode 2	Mean	SD
Subject A	4.6	4.4	117.1	140.8	128.9	16.7	30.2	30.2	30.2	0.0
Subject B	4.8	5.7	181.6	96.1	138.8	60.5	41.6	26.1	33.8	11.0
Subject C	4.6	4.6	202.5	194.4	198.5	5.7	57.6	25.4	41.5	22.8
Subject D	4.3	4.9	161.9	138.2	150.1	16.8	67.0	27.4	47.2	28.0
Subject E	5.3	4.6	222.7	222.7	222.7	0.0	25.3	25.3	25.3	0.0
Subject F	5.1	5.6	148.0	163.7	155.9	11.1	45.9	14.5	30.2	22.2
Subject G	4.6	4.7	151.7	151.7	151.7	0.0	70.9	22.4	46.6	34.3
Subject H	4.6	4.4	149.9	174.4	162.2	17.3	43.8	60.1	51.9	11.5
Subject I	4.2	5.6	115.4	74.3	94.9	29.1	25.0	25.0	25.0	0.0
Subject J	4.7	4.8	152.9	145.3	149.1	5.4	30.7	53.6	42.1	16.2
Mean	4.7	4.9	160.4	150.2			43.8	31.0		
SD	0.3	0.5	34.0	43.4			16.7	14.3		

of correlation across the group in internal versus extracted lactate is most likely to be due to skin permeability differences between human subjects and individual calibration of the device for each subject would be required for any practical application of the technique.

Discussion and conclusions

For the *in vitro* diffusion studies, cellulose ester nanoporous membrane was used to substitute for human or animal skin samples in order to avoid the large sample-to-sample differences in human skin. This allowed the electrical properties of the electrodes and the effects of the electrical waveforms on reverse iontophoresis of lactate and glucose to be studied without the nonuniformities introduced by large sample-to-sample variation in animal tissue. Moreover, the availability of excised human or animal skin samples is becoming limited for legal, ethical and safety reasons. Though excised human skin samples can be obtained, they may have a wide range of physical parameters that differ and therefore introduce a significant scatter into any skin transport experiments. Thus basic electrical characterisation studies of the electrodes and instrumentation for iontophoresis are preferably conducted in as controlled an environment as possible before skin studies commence.

In the *in vitro* glucose diffusion studies, glucose is an uncharged molecule and therefore electroosmosis is the main mechanism for glucose extraction during reverse iontophoresis. Because the efficiency of electroosmotic flow is weakly dependent upon current density in the range of 0.14–0.55 mA/cm² (Delgado-Charro and Guy 1994), a current density for iontophoresis of 0.3 mA/cm², close to

the value recommended as an upper limit by other workers (Ledger 1992), was used.

Electroosmotic flow is always in the same direction as the flow of counter ions. Since human skin is negatively charged under physiological conditions (Burnette 1987), the counter ions are cations and the electroosmotic flow is thus from anode to cathode (Guy et al 2000; Pikal 2001). The nanoporous membrane used in the diffusion cell is negatively charged. Therefore, electroosmotic flow in this model system should be from anode to cathode during direct current reverse iontophoresis and thus more glucose should be found at cathode rather than anode. However, in this model system, extracted level of glucose at the anode was slightly greater than that found at the cathode during direct current reverse iontophoresis but this was not significant. This result was not in agreement with other researchers (Rao et al 1993, 1995). The reason that the anodal extraction is greater than the cathodal extraction may be due to neutral or positively charged pores existing in the membrane (Pikal and Shah 1990). It is also possible that some masking of negative charge in the nanoporous membrane or some membrane polarization occurred which reduced the effect of cathodal electroosmosis.

As can be seen in Figure 3b, the amount of glucose extracted in the model system at either 10 or 15 minutes switching time was significantly more than that obtained by continuous current passage without switching of the electrode polarity (ie, cathode and anode held constant) but this was inconsistent with the results reported by Santi and Guy (1996). They found that total mannitol extractions were statistically equivalent for the continuous current passage with




Figure 4 Average results of long duration bipolar direct current application (current density of 0.3 mA/cm², polarity of electrodes reversed at intervals of 15 minutes, experimental time of 60 minutes) on human transdermal extraction of (a) lactate and (b) glucose (mean \pm SD; n = 10). Extraction of lactate or glucose by reverse iontophoresis was significantly higher ($p < 0.001$ for both cases) than that of the control sample.

or without switching, albeit that a different sugar molecule was studied. The increased glucose extracted in this study with electrode polarity reversal may be due to effective depolarization of the membrane (Banga and Chien 1988).

Based on the findings, it is postulated that a 15 minute time interval for electrode polarity reversal is optimum in an applied current waveform as it extracts significantly more glucose than found in the control gel (at least 3.9 times greater)

and more than DC current passage alone (at least 3.1 times greater) within the present study.

In the *in vitro* lactate diffusion studies, results showed that reverse iontophoresis could facilitate lactate extraction and applied current waveforms governed the amount of lactate to be extracted. Lactate, a negatively charged ion, is mainly extracted by electromigration during reverse iontophoresis. Based on a general principle of electricity, with like charges




Figure 5 Graphical comparison of real blood glucose levels of healthy subjects and glucose levels in the collection methylcellulose gel after reverse iontophoresis.

repelling each other and opposite charges attracting each other, lactate passed easily across the nanoporous membrane to the anode during continuous current passage. Therefore, more lactate should be detected at the anode than at the cathode.

Clearly, the time of the application of positive polarity to the electrode in any chamber during periodic electrode polarity reversal is half of that of the continuous current

conditions. However, the lactate found in either chamber in the polarity reversal mode tended to be a function of the overall positive and negative current cycles in each chamber and as such the 15 min switching period extracted a good level of lactate. It was deemed worthwhile to try and extract both lactate and glucose from the human subjects utilizing the 15 minute switching interval.




Figure 6 Graphical comparison of real blood lactate levels of healthy subjects and lactate levels in the collection methylcellulose gel after reverse iontophoresis.

In the human studies, 40% of the subjects still displayed very mild erythema even when bipolar current was employed for reverse iontophoresis despite the polarity reversal of the electrodes.

Relatively few researchers have investigated glucose extraction by reverse iontophoresis in human subjects (see for example: Rao et al 1995; Potts et al 2002; Sieg et al 2004a, 2004b) and no researchers have studied the effects of reverse iontophoresis on lactate extraction both in human subjects and *in vitro*. Rao and colleagues (1995) passed a direct current (0.25 mA/cm²) across the human skin (ventral forearm) of subjects for a period of 60 minutes to extract glucose out of the skin and they found that around 5.83 nmol of glucose could be extracted at the cathode, somewhat less than found in this study. In the present study around 14 nmol of glucose was extracted by the application of bipolar direct current for 60 minutes. The possible explanation might be that we used a slightly higher current density (0.3 mA/cm²) compared with the study by Rao and colleagues (1995). More importantly, we used a bipolar direct current rather than a direct current. Since glucose is mainly extracted out of the skin by electroosmosis and bipolar direct current can prevent the reduction in the level of electroosmosis, bipolar direct current can promote more glucose extraction than direct current.

In this study reverse iontophoresis significantly promoted transdermal lactate extraction, around 2.5 times greater than diffusion alone, but there is no other human study for data comparison.

The concentration levels of glucose and lactate found in the gel electrodes had average values >150 μM for glucose and >330 μM for lactate. These concentration levels would be eminently suitable for biosensor detection directly in the gel electrodes allowing continuous monitoring of both.

In summary, both glucose and lactate can be effectively and noninvasively extracted from human subjects simultaneously by reverse iontophoresis. This illustrates the potential for the development of a continuous and noninvasive system to concurrently monitor blood glucose and lactate levels. To arrive at this a calibration of internal versus external glucose and lactate will be required for each individual subject. We have only been able to study this in a limited manner in the current work. As shown in Figures 5 and 6, there was a poor correlation across the subject group between blood glucose and extracted glucose levels and blood lactate and extracted lactate levels. However, this is to be expected because of the known skin to skin permeability variations between human subjects. This was overcome in the GlucoWatch® and a calibration was possible for each individual user. Each user

was able to calibrate the device by utilising a single finger stick internal blood glucose sample which the device correlated with the reverse iontophoresis glucose level of the wearer at a single point in time (Tamada et al 1999). This was then used in an algorithm to link reverse iontophoresis glucose to internal glucose without the need for any further blood tests for a period of up to 12 hours. It is hoped that long term temporal studies on individual subjects will allow similar calibration procedures to be developed for lactate and glucose in simultaneous monitoring.

Acknowledgments

We would like to thank Brian Cartledge for his technical help, Canniesburn Research Trust for funding part of this work and John Swire and Sons Ltd. (HK) for the support of the studies of Congo TS CHING with a James Henry Scott Scholarship. There are no conflicts of interest to report.

References

- Abramson D, Scalea TM, Hitchcock R, et al. 1993. Lactate clearance and survival following injury. *J Trauma*, 35:584–9.
- Amos AF, McCarty DJ, Zimmet P. 1997. The rising global burden of diabetes and its complications: estimates and projections to the year 2010. *Diabet Med*, 14(Suppl 5):S1–S85.
- Artru F, Jourdan C, Perret-Liaudet A, et al. 1998. Low brain tissue oxygen pressure: incidence and corrective therapies. *Neurol Res*, 20(Suppl 1): S48–S51.
- Bakker J, Coffernils M, Leon M, et al. 1991. Blood lactate levels are superior to oxygen-derived variables in predicting outcome in human septic shock. *Chest*, 99:956–62.
- Banga AK, Chein YW. 1998. Iontophoretic delivery of drugs: Fundamentals, developments and biomedical applications. *J Control Release*, 7:1–24.
- Bell DSH. 1994. Stroke in the diabetic patient. *Diabetes Care*, 17:213–19.
- de Boer J, Pottthoff H, Mulder PO, et al. 1994. Lactate monitoring with subcutaneous microdialysis in patients with shock: a pilot study. *Circ Shock*, 43: 57–63.
- Brancati FL, Whelton PL, Randall BL, et al. 1997. Risk of end-stage renal disease in diabetes mellitus: a prospective cohort study of men screened for MRFIT. Multiple Risk Factor Intervention Trial. *JAMA*, 278:2069–74.
- Burnette RR, Ongpipattanukul B. 1987. Characterization of the perm selective properties of excised human skin during iontophoresis. *J Pharm Sci US*, 76:765–73.
- Cairns CB, Moore FA, Haenel JB, et al. 1997. Evidence for early supply independent mitochondrial dysfunction in patients developing multiple organ failure after trauma. *J Trauma*, 42:532–6.
- Carrier M, Trudelle S, Thai P, et al. 1998. Ischemic threshold during cold blood cardioplegic arrest: monitoring with tissue pH and pO₂. *J Cardiovasc Surg (Torino)*, 39:593–7.
- Chien YW, Siddiqui O, Sun Y, et al. 1987. Transdermal iontophoretic delivery of therapeutic peptides/proteins. I: Insulin. *Ann N Y Acad Sci*, 507:32–51.
- Chien YW, Lelawongs P, Siddiqui O, et al. 1990. Facilitated transdermal delivery of therapeutic peptides and proteins by iontophoretic delivery devices. *J Control Release*, 13(2–3):263–78.
- Ching CTS, Camilleri I, Connolly P. 2005. A low-cost, programmable device for versatile current delivery in iontophoresis applications. *Sensor Actuat B-Chem*, 106:534–40.

- Connolly P, Cotton C, Morin F. 2002. Opportunities at the skin interface for continuous patient monitoring: A reverse iontophoresis model tested on lactate and glucose. *IEEE Trans Nanobioscience*, 1:37–41.
- Davis BL, Kuznicki J, Praveen SS, et al. 2004. Lower-extremity amputations in patients with diabetes: pre- and post-surgical decisions related to successful rehabilitation. *Diabetes Metab Res Rev*, 20(Suppl 1): S45–S50.
- Degim IT, Ilbasnis S, Dundaroz R, et al. 2003. Reverse iontophoresis: a non-invasive technique for measuring blood urea level. *Pediatr Nephrol*, 18:1032–7.
- Delgado-Charro MB, Guy RH. 1994. Characterization of convective solvent flow during iontophoresis. *Pharm Res*, 11:929–35.
- Elia S, Liu P, Hilgenberg A, et al. 1991. Coronary haemodynamics and myocardial metabolism during weaning from mechanical ventilation in cardiac surgical patients. *Can J Anaesth*, 38:564–71.
- Evans J. 1995. Causes of blindness and partial sight in England and Wales 1990–1991. HMSO.
- Gersh BJ, Anderson JL. 1993. Thrombolysis and myocardial salvage. Results of clinical trials and the animal paradigm – paradoxical or predictable? *Circulation*, 88:296–306.
- Guy RH, Kalia YN, Delgado-Charro MB, et al. 2000. Iontophoresis: electrorepulsion and electroosmosis. *J Control Release*, 64(1–3):129–32.
- Habicht JM, Wolff T, Langemann H, et al. 1998. Intraoperative and post-operative microdialysis measurement of the human heart – feasibility and initial results. *Swiss Surg*, (Suppl 2):26–30.
- Hildingsson U, Sellden H, Ungerstedt U, et al. 1996. Microdialysis for metabolic monitoring in neonates after surgery. *Acta Paediatr*, 85:589–94.
- Hirvonen J, Hueber F, Guy RH. 1995. Current profile regulates iontophoretic delivery of amino acids across the skin. *J Control Release*, 37:239–49.
- Howard JP, Drake TR, Kellogg DL. 1995. Effects of alternating current iontophoresis on drug delivery. *Arch Phys Med Rehabil*, 76:463–6.
- [IDF] International Diabetes Federation. 2003. Diabetes Atlas 2003, 2nd edition. Brussels: International Diabetes Federation.
- King H, Rewers M. 1993. Global estimates for prevalence of diabetes mellitus and impaired glucose tolerance in adults. WHO Ad Hoc Diabetes Reporting Group. *Diabetes Care*, 16:157–77.
- Laing SP, Swerdlow AJ, Slater SD, et al. 1999a. The British Diabetic Association Cohort Study, I: all-cause mortality in patients with insulin-treated diabetes mellitus. *Diabet Med*, 16:459–65.
- Laing SP, Swerdlow AJ, Slater SD, et al. 1999b. The British Diabetic Association Cohort Study, II: cause-specific mortality in patients with insulin-treated diabetes mellitus. *Diabet Med*, 16:466–71.
- Ledger PW. 1992. Skin biological issues in electrically enhanced transdermal delivery. *Adv Drug Deliver Rev*, 9(2–3):289–307.
- Levrant J, Ciebiera JP, Chave S, et al. 1998. Mild hyperlactatemia in stable septic patients is due to impaired lactate clearance rather than overproduction. *Am J Respir Crit Care Med*, 157(4 Pt 1):1021–6.
- Marecaux G, Pinsky MR, Dupont E, et al. 1996. Blood lactate levels are better prognostic indicators than TNF and IL-6 levels in patients with septic shock. *Intens Care Med*, 22:404–8.
- Mendelowitsch A, Mergner GW, Shuaib A, et al. 1998. Cortical brain microdialysis and temperature monitoring during hypothermic circulatory arrest in humans. *J Neurol Neurosurg Psychiatry*, 64:611–18.
- Menzel M, Doppenberg EM, Zauner A, et al. 1999. Increased inspired oxygen concentration as a factor in improved brain tissue oxygenation and tissue lactate levels after severe human head injury. *J Neurosurg*, 91:1–10.
- Merino V, Kalia YN, Guy RH. 1997. Transdermal therapy and diagnosis by iontophoresis. *Trends Biotechnol*, 15:288–290.
- Merino V, Lopez A, Hochstrasser D, et al. 1999. Noninvasive sampling of phenylalanine by reverse iontophoresis. *J Control Release*, 61(1–2):65–9.
- Mize NK, Buttery M, Daddona P, et al. 1997. Reverse iontophoresis: monitoring prostaglandin E2 associated with cutaneous inflammation in vivo. *Exp Dermatol*, 6:298–302.
- Moomey Jr CB, Melton SM, Croce MA, et al. 1999. Prognostic value of blood lactate, base deficit, and oxygen-derived variables in an LD50 model of penetrating trauma. *Crit Care Med*, 27:154–61.
- Murray MJ, Gonze MD, Nowak LR, et al. 1994. Serum D(-)-lactate levels as an aid to diagnosing acute intestinal ischemia. *Am J Surg*, 167:575–8.
- Numajiri S, Sugibayashi K, Morimoto Y. 1993. Non-invasive sampling of lactic acid ions by iontophoresis using chloride ion in the body as an internal standard. *J Pharm Biomed Anal*, 11:903–9.
- Pickup J. 2003. Diabetic control and its measurements, In: Pickup JC, Williams G (Eds.). *Textbook of Diabetes*, 3rd ed. Oxford: Blackwell, pp. 34.1–34.17.
- Pikal MJ, Shah S. 1990. Transport mechanisms in iontophoresis. III. An experimental study of the contributions of electroosmotic flow and permeability change in transport of low and high molecular weight solutes. *Pharm Res*, 7:222–9.
- Pikal MJ. 2001. The role of electroosmotic flow in transdermal iontophoresis. *Adv Drug Deliv Rev*, 46(1–3):281–305.
- Pitzer KR, Desai S, Dunn T, et al. 2001. Detection of hypoglycaemia with the GlucoWatch biographer. *Diabetes Care*, 24:881–5.
- Potts RO, Tamada JA, Tierney MJ. 2002. Glucose monitoring by reverse iontophoresis. *Diabetes Metab Res Rev*, 18(Suppl 1):S49–S53.
- Rao G, Glikfeld P, Guy RH. 1993. Reverse iontophoresis: Development of a noninvasive approach for glucose monitoring. *Pharm Res*, 10:1751–5.
- Rao G, Guy RH, Glikfeld P, et al. 1995. Reverse iontophoresis: Noninvasive glucose monitoring in vivo in human. *Pharm Res*, 12:1869–73.
- Rasmussen I, Hillered L, Ungerstedt U, et al. 1994. Detection of liver ischemia using microdialysis during experimental peritonitis in pigs. *Shock*, 1:60–6.
- Rojdmark J, Blomqvist L, Malm M, et al. 1998. Metabolism in myocutaneous flaps studied by in situ microdialysis. *Scand J Plast Reconstr Surg Hand Surg*, 32:27–34.
- Santi P, Guy RH. 1996. Reverse iontophoresis – parameters determining electroosmotic flow. I. PH and ionic strength. *J Control Release*, 38(2–3):159–65.
- Sieg A, Guy RH, Delgado-Charro MB. 2004. Simultaneous extraction of urea and glucose by reverse iontophoresis in vivo. *Pharm Res*, 21:1805–10.
- Sieg A, Guy RH, Delgado-Charro MB. 2004b. Noninvasive glucose monitoring by reverse iontophoresis in vivo: Application of the internal standard concept. *Clin Chem*, 50:1383–90.
- Slomovitz BM, Lavery RF, Tortella BJ, et al. 1998. Validation of a handheld lactate device in determination of blood lactate in critically injured patients. *Crit Care Med*, 26:1523–8.
- Stephenson JM, Kenny S, Stevens LK, et al. 1995. Proteinuria and mortality in diabetes: the WHO Multinational Study of Vascular Disease in Diabetes. *Diabet Med*, 12:149–55.
- Tamada JA, Garg S, Jovanovic L, et al. 1999. Noninvasive glucose monitoring: Comprehensive clinical results. *JAMA*, 282:1839–44.
- Tierney MJ, Tamada JA, Potts RO, et al. Cygnus Research Team. 2001. Clinical evaluation of the GlucoWatch® biographer: a continual, non-invasive glucose monitor for patients with diabetes. *Biosens Bioelectron*, 16(9–12):621–9.
- The Diabetes Research in Children Network (DirecNet) Study Group. 2004. Accuracy of the GlucoWatch G2 Biographer and the continuous glucose monitoring system during hypoglycemia: Experience of the Diabetes Research in Children Network. *Diabetes Care*, 27:722–6.
- The UK Prospective Diabetes Study Group (UKPDS) Group. 1998a. Intensive blood-glucose control with sulphonylureas or insulin compared with conventional treatment and risk of complications in patients with type 2 diabetes (UKPDS 33). *Lancet*, 352:837–53.
- The UK Prospective Diabetes Study Group (UKPDS) Group. 1998b. Effect of intensive blood-glucose control with metformin on complications in overweight patients with type 2 diabetes (UKPDS 34). *Lancet*, 352:854–65.

- The UK Prospective Diabetes Study Group. 1998c. Efficacy of atenolol and captopril in reducing risk of macrovascular and microvascular complications in type 2 diabetes: UKPDS 39. *BMJ*, 317:713–20.
- The UK Prospective Diabetes Study Group. 1998d. Tight blood pressure control and risk of macrovascular and microvascular complications in type 2 diabetes: UKPDS 38. *BMJ*, 317:703–13.
- Tomohira Y, Machida Y, Onishi H, et al. 1997. Iontophoretic transdermal absorption of insulin and calcitonin in rats with newly-devised switching technique and addition of urea. *Int J Pharm*, 155:231–9.
- Valadka AB, Goodman JC, Gopinath SP, et al. 1998. Comparison of brain tissue oxygen tension to microdialysis-based measures of cerebral ischemia in fatally head-injured humans. *J Neurotrauma*, 15:509–19.
- Vincent JL. 1996. End-points of resuscitation: arterial blood pressure, oxygen delivery, blood lactate, or...? *Intensive Care Med*, 22:3–5.
- Vincent JL. 1998. Blood lactate levels: index of tissue oxygenation. *Int J Intensive Care*, 5:48–53.
- Volpe G, Moscone D, Compagnone D, et al. 1995. In vivo continuous monitoring of L-lactate coupling subcutaneous microdialysis and an electrochemical biocell. *Sensor Actuat B-Chem*, 24(1–3):138–41.
- Wang SL, Head J, Stevens L, et al. 1996. Excess mortality and its relation to hypertension and proteinuria in diabetic patients. The World Health Organization Multinational Study of Vascular Disease in Diabetes. *Diabetes Care*, 19:305–12.
- Westgren M, Kruger K, Ek S, et al. 1998. Lactate compared with pH analysis at fetal scalp blood sampling: a prospective randomised study. *Br J Obstet Gynaecol*, 105:29–33.
- Westgren M, Kublickas M, Kruger K. 1999. Role of lactate measurements during labor. *Obstet Gynecol Surv*, 54:43–8.
- Wild S, Roglic G, Green A, et al. 2004. Global Prevalence of Diabetes: Estimates for the year 2000 and projections for 2030. *Diabetes Care*, 27:1047–53.
- Zauner A, Doppenberg E, Soukup J, et al. 1998. Extended neuromonitoring: new therapeutic opportunities? *Neurol Res*, 20(Suppl 1):S85–S90.

