Comparability and repeatability of different methods of corneal astigmatism assessment

Tiago B Ferreira
Filomena J Ribeiro

Department of Ophthalmology,
Hospital da Luz, Lisbon, Portugal

Purpose: To assess the comparability and repeatability of keratometric and astigmatism values measured by four techniques: Orbscan IIz® (Bausch and Lomb), Lenstar LS 900® (Haag-Streit), Cassini® (i-Optics), and Total Cassini (anterior + posterior surface), in healthy volunteers.

Patients and methods: Fifteen healthy volunteers (30 eyes) were assessed by the four techniques. In each eye, three consecutive measures were performed by the same operator. Keratometric and astigmatism values were recorded. The intraclass correlation coefficient (ICC) was used to assess comparability and repeatability. Agreement between measurement techniques was evaluated with Bland–Altman plots.

Results: Comparability was high between all measurement techniques for minimum keratometry (K1), maximum keratometry (K2), astigmatism magnitude, and astigmatism axis, with ICC >0.900, except for astigmatism magnitude measured by Cassini compared to Lenstar (ICC =0.798) and Orbscan compared to Lenstar (ICC =0.810). However, there were some differences in the median values of K1 and K2 between measurement techniques, and the Bland–Altman plots showed a wide data spread for all variables, except for astigmatism magnitude measured by Cassini and Total Cassini. For J0 and J45, comparability was only high for J0 between Cassini and Orbscan. Repeatability was also high for all measurement techniques except for K2 (ICC =0.814) and J45 (ICC =0.621) measured by Cassini.

Conclusion: All measurement techniques showed high comparability regarding K1, K2, and astigmatism axis. Although posterior corneal surface is known to influence these measurements, comparability was high between Cassini and Total Cassini regarding astigmatism magnitude and axis. However, the wide data spread suggests that none of these devices should be used interchangeably.

Keywords: astigmatism, keratometry, topography, repeatability, comparability

Introduction

Accurate measurements of corneal power and astigmatism are crucial to attain optimal results in cataract surgery.^{1,2} Given the variety of currently available devices and the lack of a recognized gold standard, it is important to assess their comparability and repeatability, the former to evaluate the potential exchangeability of measurements obtained by different techniques and the latter to help establish which technique or instrument performs better.¹ Although several studies have previously investigated the comparability and repeatability of different available devices regarding corneal power and astigmatism,^{3–6} namely the Cassini® (i-Optics) device used in this study,^{1,2,7,8} no study has compared the corneal power and astigmatism measured using Orbscan IIz® (Bausch and Lomb), Lenstar LS 900® (Haag-Streit), and Cassini and Total Cassini (the sum of the anterior and posterior corneal surfaces) in the same eyes and by the same operator.

Correspondence: Tiago B Ferreira
Department of Ophthalmology, Hospital
da Luz, Avenida Lusitana 100, Lisboa
1500-650, Portugal
Tel +351 217 10 4400
Fax +351 217 10 4409
Email tiagoferreira@netcabo.pt

Cassini (i-Optics) is a new technology specifically developed to assess eyes before cataract surgery.⁹ This technology has been shown to have a very high specificity in estimating corneal keratometry even in post-LASIK, keratoconic, and cross-linked corneas,¹⁰ and also a high precision in reporting corneal astigmatism.¹¹

The goal of this work was to assess the comparability and repeatability of keratometric and astigmatism values measured by four techniques: Orbscan IIz, Lenstar LS 900, Cassini, and Total Cassini (anterior + posterior surface), in healthy volunteers.

Patients and methods

Setting and population sample

This was an institutional cross-sectional study that included 15 healthy volunteers (30 eyes). All eyes were assessed by the four techniques. Three consecutive measures (10 minutes apart) were performed in each eye by the same operator, using the four techniques. Keratometric and astigmatism values were recorded. Inclusion criteria were healthy individuals aged 18–50 years with a corrected visual acuity of 0.00 logMAR or better. Exclusion criteria were a history of ocular pathology, trauma, contact lens wear, systemic or local medications, and ocular surgery. In addition, patients with anterior segment pathologies such as dry eye, Meibomian gland disease, corneal disease, or abnormal topographies were excluded from this study. All participants were in the proper head positioning, and targets were positioned as instructed by the manufacturer of each device. The sequence of the measurements with the three devices was randomly chosen. The study protocol was approved by Hospital da Luz Institutional Review Board. All participants provided written informed consent.

Automated topographers

Topography data were obtained with Orbscan IIz (Bausch and Lomb), Lenstar LS 900 (Haag-Streit), and Cassini (i-Optics).

Orbscan IIz is a Placido-based multidimensional system that combines slit-scan and Placido ring technology, providing a complete analysis of the corneal surface, evaluating all corneal curvatures. Keratometry from the 3.0 mm zone was used to maximize comparability between devices. The Lenstar LS 900 uses 32 measuring points arranged in two concentric rings (outer 2.3 mm, inner 1.65 mm) of 16 measuring points each.¹² Cassini is a topographer that uses multicolor point-to-point (up to 700) ray tracing, combined with second Purkinje Imaging Technology.¹³ A detailed description of each

topographer has been previously described.⁹ Keratometry is measured in the 3.0 mm zone.

Calibration of all topographers was performed according to the manufacturer's instructions.

Measurements

Assessed parameters were minimum (K1), maximum (K2), mean keratometry (Km), astigmatism magnitude (D), and axis (°). For the Cassini, these values were recorded for the anterior corneal surface and for total corneal astigmatism (anterior + posterior surface). Vectors J0 and J45 were also assessed for the cardinal (0°–180°) and oblique (45°–135°) meridians and were calculated using the formulas $J0 = D \times \cos(2\pi \times \text{axis}/180)$ and $J45 = D \times \sin(2\pi \times \text{axis}/180)$, according to Thibos and Horner.¹⁴

Power vectors were conceived as a way of transforming conventional refractive error, or keratometric data, into mutually independent, orthogonal components, better suited to statistical analysis. Vector analysis permits a complete description of astigmatism characteristics¹⁵ and allows the comparison of both orientation and power.

Statistical analysis

After Shapiro–Wilk tests of all variables, the Wilcoxon test was used to compare measurements performed on the different pairs of devices. The intraclass correlation coefficient (ICC) was used to assess comparability and repeatability. The ICC expresses the consistency of repeated measurements, ranging from 0 to 1. An ICC <0.75 indicates poor repeatability, from 0.75 to 0.89 moderate repeatability, and >0.90 high repeatability.¹

When comparing axis, 180° was added to or subtracted from the measured axis so that measurement differences between methods were never >90°. ^{8,9,11} Analysis of agreement between each pair of devices was performed using Bland–Altman plots.^{9,16} The 95% limits of agreement (LoAs) represent the limits of the range for the 95% of differences between each pair of devices. According to the Bonferroni correction, tests were considered significant at $P < 0.008$ significance level (two-tailed). Data were processed using IBM SPSS 21 software (Armonk, NY, USA).

Results

Comparability of keratometry readings

Univariate analysis comparing K1 and K2, as assessed by Total Cassini, Cassini, Orbscan, and Lenstar, showed that there were differences in the median values of K1 and K2 between Lenstar and Orbscan (Table 1). The ICC used

Table 1 Descriptive statistics

Parameter	Total Cassini	Cassini®	Orbiscan IIz®	Lenstar LS 900®
K1 (D)		42.77 (36.93–46.06)	42.90* (36.90–45.70)	42.82* (36.86–45.76)
K2 (D)		44.31 (41.24–47.06)	44.17* (40.87–46.43)	44.06* (41.32–46.63)
Astigmatism (D)	0.88 (0.44–2.04)	0.89 (0.25–5.80)	0.80 (0.27–2.17)	0.87 (0.34–3.04)
Axis (°)	89.33 (11.33–147.50)	91.92 (10.67–167.00)	93.00 (16.00–167.00)	94.33 (12.67–163.00)
J0 (D)	0.23* (–0.92 to 0.72)	0.30* (–0.86 to 2.87)	0.33 (–0.78 to 2.82)	0.32 (–0.82 to 1.48)
J45 (D)	–0.01 (–0.76 to 0.36)	0.03* (–0.79 to 0.56)	0.04 (–0.54 to 1.63)	0.05* (–0.45 to 1.34)

Notes: All values are presented as median (minimum–maximum). Orbiscan showed higher values of K1 and K2 compared to Lenstar. Cassini showed a higher value of J0 compared to Total Cassini and a lower value of J45 compared to Lenstar. * $P < 0.005$.

Abbreviations: K1, minimum keratometry; K2, maximum keratometry.

to assess comparability showed that comparability was high between all measurement techniques for K1 and K2. Agreement between assessment methods for Km is further illustrated in the Bland–Altman plots (Figure 1). The Cassini vs Lenstar agreement regarding Km showed the closest to 0 mean difference (–0.030) but the highest range of LoA (2.397). The Orbiscan vs Lenstar agreement showed the highest mean difference (–0.143) and the lowest range of LoA (0.301).

Comparability of astigmatism evaluation

Univariate analysis comparing astigmatism magnitude, astigmatism axis, J0, and J45, as assessed by Total Cassini, Cassini, Orbiscan and Lenstar, showed that there were differences in the median values of J0 between Total Cassini and Cassini, and of J45 between Cassini and Lenstar (Table 1). The ICC used to assess comparability was high between all measurement techniques for astigmatism magnitude and astigmatism axis, with ICC > 0.900 , except for astigmatism magnitude measured by Cassini compared to Lenstar (ICC = 0.798) and Orbiscan compared to Lenstar (ICC = 0.810). For J0 and J45, comparability was only high for J0 between Cassini and Orbiscan (ICC = 0.989), with all other comparisons showing ICC < 0.900 . Although all comparisons showed a $P < 0.001$, J45 between Cassini and Orbiscan and between Orbiscan and Lenstar were notably low (ICC = 0.522 and ICC = 0.690, respectively). When comparing Cassini and Total Cassini regarding both astigmatism magnitude and axis, comparability was high (ICC = 0.941 and ICC = 0.983, respectively, $P < 0.001$).

Agreement between assessment methods for astigmatism magnitude and astigmatism axis is further illustrated in the Bland–Altman plots (Figure 1). The best agreement for astigmatism magnitude was between Total Cassini and Cassini, having the closest to 0 mean difference (0.014) and the lowest range of LoA (0.862). The agreement between Cassini and Lenstar showed the highest mean difference (0.201), and

the Orbiscan vs Lenstar the highest range of LoA (2.923). Regarding astigmatism axis, all comparisons showed a wide data spread, with the agreement between Total Cassini and Lenstar showing the closest to 0 mean difference (0.802), and the Total Cassini vs Cassini the lowest range of LoA (34.385). The agreement between Total Cassini and Cassini showed the highest mean difference (2.928), and the Total Cassini vs Orbiscan the highest range of LoA (63.564).

Repeatability

The ICC used to assess repeatability showed that this was high for all measurement techniques and assessed parameters, with ICC > 0.900 , except for K2 and J45 measured by Cassini (ICC = 0.814 and ICC = 0.621, respectively) (Table 2).

Discussion

Given the importance of evaluating the potential exchangeability of corneal power and astigmatism measurements,¹ we assessed the comparability and repeatability of keratometric and astigmatism values measured by four techniques – Orbiscan IIz, Lenstar LS 900, Cassini, and Total Cassini (anterior + posterior surface) – in 30 eyes of 15 healthy volunteers.

Comparability

The results reported in the literature concerning comparability between devices are not consensual. Ventura et al found no differences in mean values of corneal power when comparing Cassini with Lenstar, but reported a significant difference in corneal power between Cassini and a Placido-based topographer.¹ However, Klijn et al reported differences in corneal power between Cassini and Lenstar and Cassini and a Placido-based topographer, although the authors considered the differences to be of negligible clinical relevance.⁷ In addition, a comparison between Cassini and a Placido-based corneal topographer showed no differences in corneal power,² although another study comparing Cassini and a Placido-based corneal topographer showed differences

Figure 1 Bland-Altman plots for astigmatism axis, astigmatism magnitude, and Km. The limits of agreement are shown by the red line. For astigmatism axis, the best agreement was between Total Cassini and Lenstar LS 900® and the worst was between Total Cassini and Cassini®. For astigmatism magnitude, the best agreement was between Total Cassini and Cassini, and the worst was between Cassini and Lenstar. Regarding Km, the best agreement was between Cassini and Lenstar, and the worst was between Orbiscan Ilz® and Lenstar.

Abbreviation: Km, mean keratometry.

Table 2 Repeatability and comparability

Parameter	Intra-keratometer (ICC, 95% CI)			Inter-keratometer (ICC, 95% CI)		
	Cassini®	Orbscan IIz®	Lenstar LS 900®	Cassini–Orbscan	Cassini–Lenstar	Orbscan–Lenstar
K1	0.954 (0.893–0.980)	0.982 (0.959–0.992)	0.997 (0.993–0.999)	0.979 (0.954–0.990)	0.947 (0.882–0.976)	0.969 (0.932–0.986)
K2	0.814 (0.609–0.918)	0.979 (0.953–0.991)	0.997 (0.993–0.999)	0.954 (0.901–0.979)	0.949 (0.886–0.977)	0.968 (0.930–0.985)
Astigmatism	0.913 (0.806–0.963)	0.955 (0.901–0.980)	0.942 (0.859–0.977)	0.964 (0.922–0.983)	0.798 (0.550–0.910)	0.810 (0.583–0.913)
Axis	0.984 (0.962–0.993)	0.970 (0.933–0.987)	0.978 (0.945–0.991)	0.976 (0.948–0.989)	0.960 (0.911–0.982)	0.964 (0.921–0.984)
J0	0.918 (0.825–0.963)	0.984 (0.963–0.993)	0.984 (0.960–0.994)	0.989 (0.976–0.995)	0.891 (0.761–0.950)	0.893 (0.766–0.951)
J45	0.621 (0.313–0.812)	0.920 (0.827–0.965)	0.950 (0.877–0.980)	0.522 (–0.018 to 0.776)	0.815 (0.593–0.915)	0.690 (0.320–0.859)

Notes: All values are presented as median (minimum–maximum). All ICCs, both intra-keratometer and inter-keratometer, have $P < 0.001$. Repeatability was high for all measurement techniques and assessed parameters, with ICC > 0.900 , except for K2 and J45 measured by Cassini (ICC < 0.900). Comparability was also high for most parameters (ICC > 0.900), but several inter-keratometer comparisons showed ICC < 0.900 .

Abbreviations: ICC, intraclass correlation coefficient; K1, minimum keratometry; K2, maximum keratometry.

in both K1 and K2.⁸ We found no differences in the median values of K1 and K2 between Cassini and Orbscan or Lenstar, although the median values of K1 and K2 between Lenstar and Orbscan were different. Comparability was high between all assessment methods, with ICC > 0.900 .

Agreement analysis showed that the Cassini vs Lenstar agreement regarding Km showed the closest to 0 mean difference, but it also showed the highest range of LoA (2.4), which is clinically relevant, while Orbscan vs Lenstar agreement showed the highest mean difference and the lowest range of LoA. Similar or higher ranges of LoA have been reported for Cassini–Placido and Cassini–Lenstar for corneal power, leading the authors to discourage the interchangeable use of these devices.^{1,2,8}

We also evaluated the median values of astigmatism magnitude and axis between Cassini and Orbscan or Lenstar, and they showed no differences. Comparability was also high regarding astigmatism magnitude and astigmatism axis for all paired devices, except for astigmatism magnitude measured by Cassini compared to Lenstar (ICC = 0.798) and Lenstar compared to Orbscan (ICC = 0.810). Once again, the results reported in the literature are different. Ventura et al found no differences in mean values of astigmatism magnitude when comparing Cassini with Lenstar,¹ but in contrast with two studies comparing Cassini and two different Placido-based corneal topographers, both showing differences in astigmatism magnitude.^{2,8} Our results are further supported by agreement analysis, which showed that, for astigmatism magnitude, the agreement between Cassini and Lenstar showed the highest mean difference, with the agreement between the Orbscan and Lenstar showing the highest range of LoA. Similar or higher ranges of LoA have been reported for Cassini–Placido and Cassini–Lenstar for astigmatism magnitudes.^{1,2,8} Regarding astigmatism axis, comparisons showed a wide data spread.

As for J0 and J45, our results showed differences in J45 between Cassini and Lenstar, with Lenstar showing

higher values. There were also differences in the median values of J0 and J45 between Total Cassini and Cassini. However, Ventura et al found no differences in the mean values of J45 when comparing Cassini with Lenstar.¹ As for comparability, it was only high for J0 between Cassini and Orbscan (ICC = 0.989), with all other comparisons showing ICC < 0.900 . Although all comparisons showed a $P < 0.001$, J45 between Cassini and Orbscan and between Lenstar and Orbscan were notably low (ICC = 0.522 and ICC = 0.690, respectively).

These different results may be accounted for due to the different population samples, different operators, and different Placido-based devices. Given there is no gold standard device for these measurements, no conclusions can be drawn regarding which device is the most accurate. Agreement analysis showed a wide data spread, suggesting that these devices should not be used interchangeably, despite the high ICC values.

Repeatability

Repeatability was high for all measurement techniques and assessed parameters, with ICC > 0.900 , except for K2 and J45 measured by Cassini. Two studies assessing Cassini repeatability concluded that Cassini has enhanced precision, further improving astigmatism magnitude, astigmatism axis repeatability,¹¹ and keratometry,¹⁰ even in LASIK-treated, keratoconic, and cross-linked corneas.¹⁰ However, other studies have shown different results. One study concluded that, although Cassini provided highly repeatable measurements, it had a worse repeatability than Lenstar in all parameters, with Lenstar showing the best repeatability of all studied devices.¹ However, another study reported a relatively low repeatability of corneal power measurements with Cassini, but a higher repeatability of cylinder measurements compared to both Lenstar and a Placido-based topographer.⁷ In contrast, Hidalgo et al showed a good repeatability for Cassini and Placido, both for keratometry and astigmatism.⁸

A comparison between Lenstar and two Placido-based topographers showed that Lenstar and one of the Placido-based topographers showed a reasonable repeatability for corneal power, J0, and J45, while the other Placido-based topographer showed poor repeatability for J45.³ This same study showed that one of the corneal Placido-based topographers had statistically higher repeatability for corneal astigmatism and Lenstar had lower repeatability,³ but another study concluded that Lenstar showed acceptable repeatability.⁵ Other authors have reported a very high repeatability for Placido-based topographers, with ICC >0.990 for K1, K2,^{4,6} and Km.⁴ Analysis of our results shows that the best repeatability for K1 and K2 was achieved by Lenstar, and the worst for K2 and J45 measured by Cassini. All other 95% CIs show overlap.

In conclusion, the best repeatability for K1 and K2 was achieved by Lenstar. All measurement techniques also showed a high comparability regarding K1, K2, and astigmatism axis. However, the wide data spread suggests that these devices should not be used interchangeably. Further research is required to assess the validity and precision of these measurements obtained with different devices.

Acknowledgment

The authors thank Ophtec Portugal for the funding provided for this paper, which was used to access all necessary scientific bibliography.

Disclosure

The authors report no conflicts of interest in this work.

References

1. Ventura BV, Al-Mohtaseb Z, Wang L, et al. Repeatability and comparability of corneal power and corneal astigmatism obtained from a point-source color light-emitting diode topographer, a Placido-based corneal topographer, and a low-coherence reflectometer. *J Cataract Refract Surg*. 2015;41(10):2242–2250.
2. Ventura BV, Wang L, Ali SF, et al. Comparison of corneal power, astigmatism, and wavefront aberration measurements obtained by a point-source color light-emitting diode-based topographer, a Placido-disk topographer, and a combined Placido and dual Scheimpflug device. *J Cataract Refract Surg*. 2015;41(8):1658–1671.
3. Fityo S, Bühren J, Shajari M, et al. Keratometry versus total corneal refractive power: analysis of measurement repeatability with 5 different devices in normal eyes with low astigmatism. *J Cataract Refract Surg*. 2016;42(4):569–576.
4. Hua Y, Xu Z, Qiu W, et al. Precision (repeatability and reproducibility) and agreement of corneal power measurements obtained by Topcon KR-1W and iTrace. *PLoS One*. 2016;11(1):e0147086.
5. Visser N, Berendschot TT, Verbakel F, et al. Comparability and repeatability of corneal astigmatism measurements using different measurement technologies. *J Cataract Refract Surg*. 2012;38(10):1764–1770.
6. Wang Q, Savini G, Hoffer KJ, et al. A comprehensive assessment of the precision and agreement of anterior corneal power measurements obtained using 8 different devices. *PLoS One*. 2012;7(9):e45607.
7. Klijn S, Reus NJ, Sicam VA. Evaluation of keratometry with a novel Color-LED corneal topographer. *J Refract Surg*. 2015;31(4):249–256.
8. Hidalgo IR, Rozema JJ, Dhubghaill SN, et al. Repeatability and inter-device agreement for three different methods of keratometry: Placido, Scheimpflug, and color LED corneal topography. *J Refract Surg*. 2015;31(3):176–181.
9. Ferreira TB, Ribeiro FJ. A novel color-LED corneal topographer to assess astigmatism in pseudophakic eyes. *Clin Ophthalmol*. 2016;10:1521–1529.
10. Kanellopoulos AJ, Asimellis G. Color light-emitting diode reflection topography: validation of keratometric repeatability in a large sample of wide cylindrical-range corneas. *Clin Ophthalmol*. 2015;9:245–252.
11. Kanellopoulos AJ, Asimellis G. Distribution and repeatability of corneal astigmatism measurements (magnitude and axis) evaluated with color light emitting diode reflection topography. *Cornea*. 2015;34(8):937–944.
12. Lenstar LS 900® Biometer Instruction Manual, Section 4.3.2. Koeniz: Haag-Streit AG.
13. Kanellopoulos AJ, Asimellis G. Forme fruste keratoconus imaging and validation via novel multi-spot reflection topography. *Case Rep Ophthalmol*. 2013;4(3):199–209.
14. Thibos LN, Horner D. Power vector analysis of the optical outcome of refractive surgery. *J Cataract Refract Surg*. 2001;27(1):80–85.
15. Freitas Gde O, Ambrosio R Jr, Ramos I, et al. Astigmatic vector analysis of posterior corneal surface: a comparison among healthy, forme fruste, and overt keratoconic corneas. *Am J Ophthalmol*. 2016;167:65–71.
16. McAlinden C, Khadka J, Pesudovs K. Statistical methods for conducting agreement (comparison of clinical tests) and precision (repeatability or reproducibility) studies in optometry and ophthalmology. *Ophthalmic Physiol Opt*. 2011;31(4):330–338.

Clinical Ophthalmology

Publish your work in this journal

Clinical Ophthalmology is an international, peer-reviewed journal covering all subspecialties within ophthalmology. Key topics include: Optometry; Visual science; Pharmacology and drug therapy in eye diseases; Basic Sciences; Primary and Secondary eye care; Patient Safety and Quality of Care Improvements. This journal is indexed on

Submit your manuscript here: <http://www.dovepress.com/clinical-ophthalmology-journal>

Dovepress

PubMed Central and CAS, and is the official journal of The Society of Clinical Ophthalmology (SCO). The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit <http://www.dovepress.com/testimonials.php> to read real quotes from published authors.