ORIGINAL RESEARCH

Estimation of Risk Factors for Head Slippage Using a Head Clamp System. A Retrospective Study

This article was published in the following Dove Press journal: Therapeutics and Clinical Risk Management

Kazuki Sakakura^{1,2} Ayataka Fujimoto ^[b] Naoki Ichikawa¹ Eiichi Ishikawa² Akira Matsumura² Hideo Enoki¹ Tohru Okanishi ^[b]

¹Comprehensive Epilepsy Center, Seirei Hamamatsu General Hospital, Hamamatsu, Shizuoka, Japan; ²Department of Neurosurgery, University of Tsukuba, Tsukuba, Ibaraki, Japan

Correspondence: Ayataka Fujimoto Seirei Hamamatsu General Hospital, 2-12-12 Sumiyoshi, Nakaku, Hamamatsu, Shizuoka 430-8558, Japan Tel +81-53-474-2222 Fax +81-53-475-7596 Email afujimotoscienceacademy@gmail.com



Background: Although complications have been associated with head clamp systems, few reports have described head slippage. The present study aimed to determine risk factors for head slippage and speculated that the position of head holder pins might be associated.

Patients and Methods: We reviewed medical records and compared the positions of the pinned heads of patients on fused preoperative and postoperative computerized tomography (CT) images. We measured the distance between corresponding head pins to determine head slippage. Age, sex, body weight, body mass index, surgical position, surgical duration, craniotomy volume, and the relationship between head pins and the nasion-inion (NI) line were statistically compared between patients with and without head slippage.

Results: Head slippage in 3 (10%) of 28 patients was significantly associated with the most caudal pin position (p < 0.001) and craniotomy volume (p = 0.036). Receiver operator characteristics curves indicated a cutoff of 4.5 cm from the NI line (sensitivity and specificity, 1.000 and 0.800, respectively).

Conclusion: Clamped heads can slip during surgical procedures. We found that one head pin should be located within 4.5 cm from the NI line to avoid head slippage.

Keywords: head movement, mayfield head clamp, intraoperative computed tomography, nasion-inion line, pinning

Introduction

Holders are used to immobilize the head during many neurosurgical procedures.¹ Even if their immobilization ability is not guaranteed, they must be trusted during surgery for various periods. However, slight head movements do occur during surgical procedures. Not all neurosurgeons feel absolutely comfortable with pinned head fixation. In fact, the US Food and Drug Administration in 2016 urged caution with neurosurgical head holders due to slippage, based on reports of 1000 slippages with 700 injuries.² Some reports have associated skull hematomas,³ fractures,^{4,5} air embolism,⁶ and asystole⁷ as complications of pinned head holders. However, the reasons for slippage have not been addressed and few reports have described head slippage, despite 700 associated injuries. This is probably due to difficulties with comparing pre-and post-operative head holder positions. We postulated that the pin position of the head holder might be associated with slippage. We, therefore, compared preoperative and postoperative computed tomography (CT) images of pinned head positions to determine risk factors associated with slippage.

Therapeutics and Clinical Risk Management 2020:16 189-194

189

© 2020 Sakakura et al. This work is publiched and licensed by Dove Medical Press Limited. The full terms of this license are available at https://www.dovepress.com/rerms. work you hereby accept the Terms. Non-commercial uses of the work are permitted without any further permission from Dove Medical Press Limited, provided the work is properly attributed. For permission for commercial use of this work, please see paragraphs 4.2 and 5 of our Terms (https://www.dovepress.com/rems.php).

Patients and Methods Study Design and Ethics Approval

The Ethics Committee at Seirei Hamamatsu General Hospital approved the protocol of this retrospective, crosssectional, observational, non-randomized study, which was designed in accordance with the principles of the Declaration of Helsinki (7th revision 2013). Participants were identified via a retrospective electronic chart review of patients treated between October 2015 and March 2019 at the Comprehensive Epilepsy Center, Seirei Hamamatsu General Hospital. Written, informed consent was obtained from all patients to publish their innominate data.

Patients

We reviewed the medical records of patients who underwent preoperative and postoperative CT in an operating room and fulfilled the following criteria: all CT images were acquired with the head clamped using a radiolucent carbon Mayfield three-pin head holder (Infinity XR2 Radiolucent System; Integra LifeSciences Services, Saint-Priest, France); the position of the head holder was not deliberately changed during procedures and the CT slice thickness was <1 mm. The heads of patients who weighed >50 or <50 kg were clamped with 60 and 40 lb of force, respectively.

Intraoperative CT

An internationally standardized point in the SOMATOM Definition AS64 Open scanner (Siemens AG, Munich, Germany) called the International Organization of Standardization (ISO) center represents the absolute reference point for neuronavigation. The camera of the neuronavigation system also detected the ISO center during CT, as well as a patient reference array, so head positions were automatically determined from CT images. Spatial errors theoretically do not exist between these structures.⁸ We applied this system without surface matching, using a touchless laser pointer for patient registration, and thus accurately measured head slippage.

Measurement of Head Slippage as a Primary Outcome

We used the iPlan station of Cranial Surgical Planning software (Brainlab, AG, Munich, Germany) to fuse preoperative and postoperative CT images. We adjusted all angles of CT image slices to the nasion-inion (NI) line on the iPlan station. The thickness of each slice increased from 1 to 2.5 mm using this modification and they were automatically fused using an algorithm.^{9,10} After fusion, we measured the distance between the corresponding head pins and the height of the pins from the NI-plane. We referred to the maximal distance of pins that had slipped >2.5 mm among the three pins because distances <2.4 mm could not be measured due to the thickness of the CT slices.

Other Factors as Secondary Outcomes

We assessed age, sex, body weight (BW), body mass index (BMI), surgical position, surgical duration, and craniotomy volume, as well as the relationship between each head pin and the NI line of each patient.

Surgical duration was taken as the amount of elapsed time between acquisition of the first preoperative and the first postoperative iCT images with head clamped in the holder. We did not determine the period between skin incision and wound closure. Craniotomy volume was measured using the iPlan station.

Statistical Analysis

Age, sex, BW, BMI, and the craniotomy volume were assessed using univariate regression analyses to determine the influence of variables on slippage rates. Slippage rates and the sample sizes were too small for multivariate logistic regression analysis; therefore, we repeated univariate regression analysis as appropriate. The cutoff value was calculated from receiver operating characteristics (ROC)

Table I Clinical Information

Age at surgery: median (range)	32 (2–77)
Gender (male: female)	20:8
Surgery	Electrode implantation 13 Transnasal tumor removal 5 Tumor biopsy 4 Tumor removal 3 Epileptical focus resection 1 Epileptical disconnection 1 Intracranial hematoma removal 1
Body weight (kg): median (range)	63.2 (12.8–101.4)
BMI (kg/m2): average (range)	21.71 (14.83–35.09)
Surgical position (supine: prone)	26:2
Duration between CTs (min): average (range)	184.68 (23-423)
Volume of craniotomy (cm3): median (range)	0.31 (0–63.45)

Abbreviations: BMT, Body mass index; CT, computed tomography.

Table :	2	Slippage	Cases
---------	---	----------	-------

	Age	Gender	Surgery	Distance of Slippage(mm)
Patient I	59	Male	Transnasal tumor removal	10
Patient 2	23	Male	Epileptical disconnection	2.5
Patient 3	64	Male	Tumor biopsy	2.5

curves. Statistical significance was taken at p < 0.05. All data were statistically analyzed using Sigma Plot version 14 (Systat Software, San Jose, CA, USA).

Results

Clinical Information

Among 29 patients (male, n = 21; female, n = 8) who fulfilled the inclusion criteria, one was excluded to avoid bias because CT images from this patient could be manually fused, but not be automatically by the iPlan station. Table 1 describes the clinical information.

Head Slippage: Primary Outcome

Table 2 shows that the primary outcome of head slippage occurred in 3 (10%) patients. The distance between preoperative and postoperative slipped pins ranged from 2.5 to 10 mm. However, no complications were associated with head slippage in these patients.

Representative Case I

A 59-year-old male with a pituitary tumor underwent preoperative iCT for neuronavigation registration after general anesthesia. Complications associated with tumor removal and any other causes were evaluated by



Figure I Measurement of slippage distance. Preoperative red (A) and postoperative green (B) dots on coronal images show pin slippage. The fused image by the iPlan station (C) also shows slippage. All distances between preoperative and postoperative pin positions were measured in this manner.



Figure 2 Position of most caudal head pins in representative patients with and without slippage. (A and B) Patient without slippage. Most caudal pin (arrowhead) is located near the nasion-inion line (dotted line). (C and D) Patient with slippage. Most caudal pin (arrowhead) is located >4.5 cm vertically (black line) from nasion-inion line (dotted line).

postoperative iCT. Preoperative and postoperative iCT images merged using iPlan showed that his head had slipped during the procedure (Figure 1).

Other Factors: Secondary Outcomes Relationship Between Head Pins and NI Line

The most caudal head pin position was significantly associated with head slippage (p < 0.001; Figure 2) The ROC [Please consider that "ROC has already been defined" in page 7. It does not need to be defined again in the text.] curves indicated a cutoff of 4.5 cm (sensitivity and specificity, 1.000 and 0.800, respectively).

Risk Factors for Slippage

Table 3 shows that craniotomy volume was a significant risk factor for head slippage (p = 0.036), whereas age, sex, BW, BMI, surgical position and surgical duration were not.

The ROC curves indicated a cutoff of 1.54 cm^3 (sensitivity and specificity, 0.67 and 0.72, respectively).

Table	3	Risk	Factors	for	Sliddage
labic	•	1,0010	1 accorb		Subbarge

Factors	Slippage cases (N=3)	Non Slippage Cases(N=25)	Р
Age at surgery: median (range)[Std Dev]	59 (23–64)[22.4]	29 (2–77)[21.9]	0.287
Gender (male: female)	Male n = 3, female n = 0	Male n = 17, female n = 8	0.12
Body weight (kg): median (range)[Std Dev]	63.2(63.2–70.9)[4.03]	53.45(12.8-101.4)[23.6]	0.477
BMI (kg/m2): median (range)[Std Dev]	21.36(21.36–24.17)[1.61]	21.28(14.83-35.09)[4.75]	0.477
Surgical position (supine: prone)	Supine $n = 2$, prone $n = 1$	Supine n = 24, prone n = I	0.96
Duration between pre and postoperative CTs (min): median (range)[Std Dev]	259(97-423)[165.1]	173.5 (23–345)[74.4]	0.225
Olume of craniotomy (cm3): median (range)[Std Dev]	1.356(0–63.448)[36.2]	0.454(0-36.396)[8.68]	0.036*

Note: *p<0.05.

Abbreviation: Std Dev, standard deviation.

Discussion Head Slippage: Primary Outcome

We found that heads clamped with pins slipped at a frequency of 10% during surgery. We then statistically evaluated risk factors to eliminate such slippage. The most caudal pin above 4.5 cm from NI line was significant ($p \le 0.001$) with 1.000 sensitivity and 0.800 specificity. This means that one of the three pins should be located within 4.5 cm from the NI line to stabilize the clamp. A previous study found that the center of gravity is located at the middle of the NI line.¹¹ This means that the head could slip more easily if the pins are inserted more superiorly, because they would be located further away from the center of gravity of the head.

The positioning of the Mayfield headrest represents the first surgical procedure in the operating room; newer neurosurgeons need to be educated about not only surgical skills,¹² but also correct technique and the risks associated with intraoperative slippage.

Craniotomy Volume

A larger craniotomy volume was a risk factor for head slippage. A larger craniotomy often does not achieve satisfactory cosmetic results due to shifting.¹³ However, this outcome might be statistically paradoxical due to lower sensitivity and specificity than the cutoff of the NI-line. Nonetheless, this factor should be considered before patients with larger craniotomies are scheduled to undergo procedures that require immobilizing the head.

Study Limitations

The number of participants in this preliminary retrospective study at a single facility was small. The thickness of the axial image slices had to be increased from 1 to 2.5 mm to measure distances between the head pins. We should be able to show more details with thinner slices A prospective, multicenter trial with more participants and thinner slices is needed to confirm the present findings.

Conclusions

Clamped heads can slip during surgery. One head pin should be located within 4.5 cm of the NI line to avoid head slippage.

Highlights

- · Measurement of head slippage with head clamp system
- Frequency of head slippage is 10%
- Head pin position is a risk factor for head slippage

Abbreviations

BMI, body mass index; BW, body weight; CT, computed tomography; ISO, International Organization of Standardization; NI, nasion-inion; ROC, receiver operating characteristics.

Acknowledgments

We are grateful to Dr. C. Inenaga and Dr. T. Tanaka for help with surgical procedures.

Author Contributions

All authors contributed to data analysis, drafting or revising the article, gave final approval of the version to be published, and agree to be accountable for all aspects of the work.

Funding

No funding was received for this investigation.

Disclosure

None of the authors has any conflicts of interest to declare.

References

- Goel A, Muzumdar DP, Phalke U. Head holder for microsurgical dissection. J Clin Neurosci. 2004;11(3):300–301. doi:10.1016/j. jocn.2003.09.009
- Atteya MME, Raslan S, Elkallaf M, Soliman MS, AlQalla MA. A novel method for safe and accurate placement of the rocker pins of head immobilization devices utilizing a digital caliber phantom: technical note. *Childs Nerv Syst.* 2019;35(9):1599–1602. doi:10.1007/s00381-019-04216-x
- Baerts WD, de Lange JJ, Booij LH, Broere G. Complications of the mayfield skull clamp. *Anesthesiology*. 1984;61(4):460–461. doi:10.1097/ 00000542-198410000-00017
- Matouk CC, Ellis MJ, Kalia SK, Ginsberg HJ. Skull fracture secondary to application of a mayfield skull clamp in an adult patient: case report and review of the literature. *Clin Neurol Neurosurg.* 2012;114 (6):776–778. doi:10.1016/j.clineuro.2011.12.036
- Zaazoue MA, Bedewy M, Goumnerova LC. Complications of head immobilization devices in children: contact mechanics, and analysis of a single institutional experience. *Neurosurgery*. 2018;82(5):678–685. doi:10.1093/neuros/nyx315
- Grinberg F, Slaughter TF, McGrath BJ. Probable venous air embolism associated with removal of the mayfield skull clamp. *Anesth Analg.* 1995;80(5):1049–1050. doi:10.1097/00000539-199505000-00036
- Miyoshi H, Nakamura R, Hamada H. Asystole following skull clamp to Chiari malformation. J Anesth. 2015;29(2):317. doi:10.1007/ s00540-014-1890-y
- Fujimoto A, Okanishi T, Kanai S, Sato K, Nishimura M, Enoki H. Neuronavigation-guided Frameless Stereoelectroencephalography (SEEG). *Neurol Med Chir (Tokyo)*. 2017;57(9):496–502. doi:10.2176/ nmc.tn.2017-0110
- 9. Riva M, Hiepe P, Frommert M, et al. Intraoperative computed tomography and finite element modelling for multimodal image fusion in brain surgery. *Oper Neurosurg (Hagerstown)*. 2019. doi:10.1093/ons/opz196
- Jin JY, Yin FF, Tenn SE, Medin PM, Solberg TD. Use of the BrainLAB ExacTrac X-ray 6D system in image-guided radiotherapy. *Med Dosim*. 2008;33(2):124–134. doi:10.1016/j.meddos.2008.02.005

193

- Vital JM, Senegas J. Anatomical bases of the study of the constraints to which the cervical spine is subject in the sagittal plane. A study of the center of gravity of the head. *Surg Radiol Anat.* 1986;8 (3):169–173. doi:10.1007/BF02427845
- Lemole GM Jr., Banerjee PP, Luciano C, Neckrysh S, Charbel FT. Virtual reality in neurosurgical education: part-task ventriculostomy simulation with dynamic visual and haptic feedback. *Neurosurgery*. 2007;61(1):142–149. doi:10.1227/01.neu.0000279734.22931.21
- Unterhofer C, Wipplinger C, Verius M, Recheis W, Thome C, Ortler M. Reconstruction of large cranial defects with poly-methyl-methacrylate (PMMA) using a rapid prototyping model and a new technique for intraoperative implant modeling. *Neurol Neurochir Pol.* 2017;51(3):214–220. doi:10.1016/j. pjnns.2017.02.007

Therapeutics and Clinical Risk Management

Dovepress

Publish your work in this journal

Therapeutics and Clinical Risk Management is an international, peerreviewed journal of clinical therapeutics and risk management, focusing on concise rapid reporting of clinical studies in all therapeutic areas, outcomes, safety, and programs for the effective, safe, and sustained use of medicines. This journal is indexed on PubMed Central, CAS, EMBase, Scopus and the Elsevier Bibliographic databases. The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit http://www.dovepress.com/testimonials.php to read real quotes from published authors.

Submit your manuscript here: https://www.dovepress.com/therapeutics-and-clinical-risk-management-journal