Open Access Full Text Article

# Highly Sensitive Lanthanide-Doped Nanoparticles-Based Point-of-Care Diagnosis of Human Cardiac Troponin I

Lu Chen<sup>1</sup>, Shan-Yong Zhou<sup>2</sup>, Wei Zhu<sup>3</sup>, Sheng-Ping Liu<sup>4</sup>, Jing-Xi Zhang<sup>5</sup>, He Zhuang<sup>5</sup>, Jing-Ling Zhang<sup>5</sup>, Yong-Sheng Li<sup>6</sup>, Fei Gao<sup>4</sup>

<sup>1</sup>Department of paediatrics, Fujian Maternity and Child Health Hospital, Fuzhou, 350000, People's Republic of China; <sup>2</sup>Fujian Key Laboratory of Nanomaterials, Fujian Institute of Research on the Structure of Matter, Chinese Academy of Sciences, Fuzhou, 350002, People's Republic of China; <sup>3</sup>Department of Urology, Fujian Provincial Hospital, Shengli Clinical Medical College of Fujian Medical University, Fuzhou, 350001, People's Republic of China; <sup>4</sup>Fujian Institute of Hematology, Fujian Provincial Key Laboratory on Hematology, Fujian Medical University Union Hospital, Fuzhou, 350001, People's Republic of China; <sup>5</sup>Department of Clinical Laboratory, Fujian Medical University Union Hospital, Fuzhou, 350001, People's Republic of China; <sup>6</sup>Department of Urology, Fujian Medical University Union Hospital, Fuzhou, 350001, People's Republic of China;

Correspondence: Fei Gao; Yongsheng Li, Tel/Fax +86 591-83357896-8242, Email xuzhougaofei@126.com; Lysdr163@163.com

**Introduction:** Cardiac troponin I (cTnI) has been regarded as a gold standard for early diagnosis and prognosis monitoring of acute myocardial infarction (AMI) in clinical practice. Owing to its low concentration in blood, accurate determination of cTnI often requires high sensitivity. However, current established point-of-care (POC) assays are insufficient to meet clinically analytical requirements due to their low sensitivity.

**Methods:** To this end, we established a highly sensitive and reliable POC lateral flow strip based on lanthanide-doped nanoparticles (NPs) for cTnI determination in human blood samples. The capture of cTnI on the lateral flow strip was performed in a sandwich assay, where  $Eu^{3+}$ -doped vanadate nanoparticles (GdVO<sub>4</sub>:30% Eu NPs) were used as luminescent probes to allow quantification.

**Results:** Our platform realized the analytical sensitivity enhancement with limit-of-detection (LOD) as low as 17 pg mL<sup>-1</sup> for cTnI detection, which was lower than the commercial counterpart; meanwhile, it displayed high specificity, excellent reproducibility and outstanding accuracy for analyzing clinical serum samples.

**Conclusion:** Overall, this strategy provided an ultrasensitive, cost-effective and user-friendly platform for on-site cTnI detection, demonstrating the prospect of lanthanide-doped NPs-based POC diagnosis of disease-related biomarkers.

Keywords: cardiac troponin I, GdVO<sub>4</sub>:30% Eu NPs, lateral flow test strip, point-of-care test

#### Introduction

Acute myocardial infarction (AMI), generally known as heart attack, is caused by prolonged ischemia and has the characteristics of rapid onset, poor prognosis and high fatality rate.<sup>1</sup> Regrettably, it is estimated 23.3 million people worldwide, by 2030, will die annually from cardiovascular disease.<sup>2</sup> Traditionally, AMI diagnoses mainly rely on patient symptoms, medical history, electrocardiography and medical imaging, which are time-consuming and expensive, and the diagnosis is equivocal. In this regard, the sensitive biomarker appraisal of patients with symptoms suggestive of AMI is urgently needed. According to the 4th edition definition of acute myocardial infarction published in 2018, cardiac troponin I (cTnI) is a significant indicator for AMI, owing to its high clinical sensitivity and specificity as well as prolonged release, in that cTnI remains in the bloodstream for more than 10 days and reaches to peak at ca. 1–2 days.<sup>3</sup> As the recommended clinical cut-off of cTnI is as low as 0.03 ng mL<sup>-1,4</sup> many commercial kits experience the limitation of low sensitivity. To this end, the affordable, rapid, simple, sensitive and accurate appraisal of cTnI is of great importance to early diagnosis and prognosis of AMI.<sup>5</sup>

Currently, the role of point-of-care testing (POCT) has been increasingly recognized, since analytical tests can be performed near the patients by both medical professionals and the patients themselves to facilitate health management.<sup>6,7</sup>

635

Enormous efforts have been devoted to establish POCT platforms, whose remarkable advantages lie in portability, speediness and straightforward readout.<sup>8</sup> Among the reported POCT devices, lateral flow assay (LFA), paper-based simplified sensor format, is the most prevalent means due to one-step analysis and user-friendly operation.<sup>9,10</sup> The analysis can qualitatively or quantitatively determine the level of cTnI in body fluid, typically involving conjugation of the target analyte with the probe to form a complex, which generates a readable signal. Such rapid readout can contribute to well-informed clinical decisions and allow for sample-to-answer analysis. However, one drawback is that LFA tests are subjected to poor detection limits compared with those of laboratory tests<sup>11,12</sup> and thus are insufficient to meet the clinically analytical requirements. For example, Natarajan et al.<sup>13</sup> developed a paper-based LFA platform for the quantitative determination of cTnI with limit of detection (LOD) in the range 1.28–1.40 ng mL<sup>-1</sup>.

Fluorescence-based LFA is an ideal candidate not only due to its inherently high sensitivity,<sup>14</sup> but also because it has good compatibility with the clinically analytical platforms. For LFA sensitivity enhancement, the development of highperformance luminescent nanoprobes is increasingly crucial. Trivalent lanthanide  $(Ln^{3+})$  ions in solids can emit photons efficiently in the spectral region from ultraviolet to visible or infrared due to its distinct electronic configuration. Composed of stable inorganic host and Ln<sup>3+</sup> emitters embedded in the host lattice, Ln<sup>3+</sup> ions-doped nanoparticles (NPs) typically feature unique properties, such as long photoluminescence lifetime and narrow emission bands, in addition to low-toxicity, photostability, high thermal and chemical stability, when compared with conventional molecular probes (such as organic dves and lanthanide chelates).<sup>15,16</sup> Furthermore, Ln<sup>3+</sup> ions-doped NPs based LFA can achieve highly sensitive bioanalysis. For example, Liu et al.<sup>17</sup> employed near infrared lanthanide-doped NPs modified FLA platform with low interference and low detection limit. Chen et al.<sup>18</sup> reported a lanthanide-doped polystyrene NPs based LFA, and the result was consistent with clinical diagnosis. Another feature is that an antenna needs to be introduced, which can effectively harvest the incident light and sensitize the Ln<sup>3+</sup> luminescence.<sup>19,20</sup> Different energy transfer schemes from the inorganic matrix and the dopant ions have been employed to enhance the luminescence of lanthanidedoped NPs.<sup>21</sup> Vanadate is considered as an excellent optical entity to enhance the global luminescent emission because of its large absorption cross-sections and effective substrate sensitization band.<sup>22,23</sup> Moreover, Eu<sup>3+</sup> ions in vanadate matrix have been reported to emit strong and stable luminescence.<sup>24</sup> The vanadate NPs doped with  $Eu^{3+}$  yielded the typical red luminescence of Eu<sup>3+</sup>, which can be excited with near-UV light through an energy transfer from the vanadate anion. Lanthanide-doped vanadate NPs have been intensively used in magnetic resonance imaging (MRI),<sup>25,26</sup> however, they are still in the infancy stage when applied in bioassay fields, especially in POC diagnosis,

In this work, a facile and green preparation method was developed to synthesize  $Eu^{3+}$  doped vanadate (GdVO<sub>4</sub>:30% Eu) NPs as the biological fluorescent probes to establish a LFA method (Scheme 1). Owing to the high labeling ratio as well as background-free signal, the developed LFA strip is capable of sensitively, selectively and accurately determining cTnI in human blood, and has significant clinical application prospect due to its merit of low detection limit (17



Scheme I Schematic illustration for configuration and principle of the cTnl determination.

pg mL<sup>-1</sup>), which is much lower than that of the commercial POCT kit (LOD > 100 pg mL<sup>-1</sup>). Moreover, the resulting system is user friendly due to convenient operation, visual fluorescence detection and rapid analysis. Taken together, the lanthanide-doped NPs integrated LFA platform can be a reliable assay for on-site cTnI determination, which is suitable for early screening of patients with AMI.

# **Materials and Methods**

### Materials and Apparatus

Materials and apparatus can be referred in Supplementary Information.

# Synthesis of Eu<sup>3+</sup> Doped Vanadate NPs

The GdVO<sub>4</sub> NPs doped with  $Eu^{3+}$  ( $Eu^{3+} = 2.5$ , 5, 10, 20 and 30 mol%) were prepared by a simple and facile coprecipitation method. In a typical synthesis of GdVO<sub>4</sub>:30% Eu NPs, 0.35 mmol Gd (NO<sub>3</sub>)<sub>3</sub>, 0.15 mmol Eu (NO<sub>3</sub>)<sub>3</sub>, 1–2.5 mmol sodium citrate and 0.5 mmol Na<sub>3</sub>VO<sub>3</sub> were added to 50 mL round-bottomed flask and dissolved in 20 mL distilled water. Subsequently, the reaction was conducted at 85 °C for 24 h under nitrogen atmosphere. After cooling down to room temperature, the resulting product was collected by centrifugation with speed of 10,000 rpm, washed with distilled water three times and dried to obtain the GdVO<sub>4</sub>:Eu NPs.

## Labeling of Anti-cTnI Antibodies with GdVO<sub>4</sub>:30% Eu NPs

The antibodies were labeled with GdVO<sub>4</sub>:30% Eu NPs via the classical covalent linkage method using 1-ethyl-3-(3-dimethylaminopropyl)-carbodiimide hydrochloride (EDC) and N-hydroxy-succinimide (NHS). In brief, 1% solid content of GdVO<sub>4</sub>:30% Eu NPs was first subjected to ultrasonic dispersion for 2 min, and then 200  $\mu$ L of solution was taken and centrifugated (14000 rpm, 15 min), and finally re-suspended by 1 mL of 2-(N-morpholino) ethanesulfonic acid (MES) buffer solution (50 mM, pH 5–6). Then, 50  $\mu$ L of EDC (100 mg mL<sup>-1</sup>) was added to the above solution and incubated for 5 min, and 150  $\mu$ L of NHS (100 mg mL<sup>-1</sup>) was then added and incubated for 15 min. After activation, 1 mL of MES buffer solution was added to re-suspend GdVO<sub>4</sub>:30% Eu NPs twice. Subsequently, 100–200  $\mu$ g of primary antibody against the cTnI was added to the activated GdVO<sub>4</sub>:30% Eu NPs, followed by cross-coupling reaction for 2 h. 500  $\mu$ L of blocking buffer (containing 50 mM Tris-HCl, 0.5% BSA and 0.1% proclin) was added and reacted for 1 h. Then, the target NPs-labeled antibodies were collected by centrifugation (14,000 rpm, 15 min) and washed twice with Tris buffer solution (50 mM, 1% NaCl, 0.5% BSA, 0.1% tween-20, 0.1% proclin, pH = 8). Finally, the resulting product was re-suspended by 200  $\mu$ L of Tris buffer solution.

## Fabrication of LFA Strip for cTnl Determination

The lateral flow test strip consisted of a sample binding pad, nitrocellulose (NC) membrane and absorbent pad, which were assembled sequentially onto the polyvinyl chloride (PVC) backing for supporting all the components. The sample pad was prepared using glass fiber which was pretreated by PBS containing 2% NaCl, 2% BSA, 0.5% casein, 0.1% tween-20, 0.5% surfactant S9 and 5% sucrose, and then loaded with GdVO<sub>4</sub>:30% Eu NPs labeled primary antibody against the cTnI. The sample binding pad and absorbent pad were attached on both ends of the NC membrane respectively to form a test area where cTnI antibody of another epitope (1 mg mL<sup>-1</sup>) and goat anti-mouse IgG antibody (1 mg mL<sup>-1</sup>) were immobilized in parallel to form a test line and control line, respectively. The time-resolved fluorescence immunochromatographic strip with a width of 4 mm was finally obtained and stored in sealed bag at 4 °C for further use.

## The cTnl Determination Based on the LFA Platform

The cTnI detection was conducted in a double-antibody sandwich format. 75  $\mu$ L of sample solution containing cTnI with known concentration was dropped onto the sample pad and migrated along the strip due to capillary force. After appropriate reaction time, the test strip was subjected to the fluorescent reader under the time-resolved detection mode, followed by recording fluorescent intensity ratio of test line to control line (abbreviated as T/C ratio) against cTnI

concentration. A calibration curve can be established to quantitatively indicate the cTnI concentration in real serum sample. Each experiment was repeated three times, and the data were presented as mean standard deviation.

## The cTnl Determination Based on Commercial Kit

In order to measure cTnI standard and clinical samples, 50  $\mu$ L of sample solution and 80  $\mu$ L of H<sub>3</sub>NO (10 mM) were added into the test strip involving dried HAuCl<sub>4</sub> (30 mM) on a reagent pad. After 20 min of assay time, a smartphonebased reader was used to capture the test strip images, running Image J software to analyze the signal values. Then, serially diluting cTnI antigens with serum-free cTnI were employed to prepare different concentrations of standard cTnI samples. All experiments involving blood samples (n = 86) were approved by the Ethics Committee of Fujian Medical University Union Hospital (Approval No. 2020WSJK032) and all donors signed the informed consent, in accordance with the Declaration of Helsinki. After collection, these clinical samples were measured by AFS-1000 dry-type fluorescence immunoassay analyzer. All clinical samples and cTnI antigens were stored at -80 °C until used.

To appraise the stability of the reagent used for Au-ion amplification, 30 mM HAuCl<sub>4</sub> that was dried on reagent pad followed by being vacuum-sealed and 10 mM H<sub>3</sub>NO solution were stored at 44 °C for 80 days, so as to accelerate test for color stability. Following which Au-ion amplification was performed on a standard test strip by treatment with  $0.5 \times 40$  nm Au NPs-antibodies (Ab) conjugates on test or control line. To evaluate batch reproducibility, three batches of test strips assembled with a 3.0 mm PVA barrier and Au-ion amplification setup were constructed. Additionally, in serum sample, each batch of strips was measured with three varying concentrations, namely 10,  $10^3$  and  $10^5$  pg mL<sup>-1</sup> of standard cTnI. Then smartphone-based reader and Image J software were respectively employed to capture the images and analyse the signals. The stability test of the reagent obtained on the day of preparation was original activity (100%). The relative activity for each day was calculated by dividing the corresponding test result by the level of the original activity.

## Statistic Analysis

The two-sample *t*-test was performed for evaluation of the differences in measured intensity in the cTnI clinical samples (10–40 pg mL<sup>-1</sup>), and all data are presented as the mean  $\pm$  standard deviation (SD) of at least three measurements. Experimental replicates are exhibited in respective figure captions. Statistical significance was set at P < 0.05.

# **Results and Discussion**

## Principle of the LFA Platform

The detection of cTnI on the lateral flow test strip was performed in a sandwich assay format based on the antigenantibody reaction. As illustrated in Scheme 1, primary antibody (Ab1) against the cTnI was labeled with  $GdVO_4$ :30% Eu NPs to form NPs-Ab1 conjugates which were pre-fixed on the conjugate pad. Another antibody (Ab2) against a different epitope of the cTnI antigen and goat anti-mouse IgG were immobilized on the NC membrane as test line and control line, respectively. When the liquid sample was introduced onto the sample pad, cTnI antigens moved along the test strip due to capillary force, and then they were bound with NPs-Ab1 conjugates. The formed complexes migrated to the NC membrane and were first captured by test line (T line), forming a sandwich structure. The excess NPs-Ab1 conjugates continued to migrate along the NC membrane and were finally captured in the control line (C line) by reacting with goat anti-mouse antibodies (IgG). Thus, positive and negative results could be determined by observing the fluorescent appearance in the T line and C line. The appearance of two colored lines suggests a positive result, while a negative test sample is indicated by the appearance of only the C line. One significant advantage of ratiometric quantification is its insensitivity to variation of different strips, enabling a robust readout between intra- and inter-assays. After appropriate reaction time, the strip was subjected to fluorescence detection, and a linear calibration curve can be established by plotting the intensity ratio of T/C against cTnI with known concentration, which can then be used to quantitatively indicate cTnI concentration in the test solution.

#### Characterization of the Fluorescent Probes

In this work, water-soluble GdVO<sub>4</sub>:30% Eu NPs were synthesized by a simple coprecipitation route. The structure of resulting product was identified by XRD analysis, as shown in Figure S1. All the diffraction peaks appeared to match well with pure GdVO<sub>4</sub> (JCPDS: 17–0260) and indicated the sample crystallized in tetragonal phase (space group:  $I4_1/$  amd).<sup>27,28</sup> No other phase was observed, indicating the high phase purity of the Eu-doped GdVO<sub>4</sub>. The enlarged d values (Table S1) of three strong lines can be attributed to the introduction of Eu<sup>3+</sup>,<sup>29</sup> because ionic radius of the Eu<sup>3+</sup> ion (0.947Å) is slightly larger than that of Gd<sup>3+</sup> ion (0.938Å),<sup>16</sup> thus suggesting that the Eu<sup>3+</sup> ions had been successfully incorporated into the host (GdVO<sub>4</sub>) crystal lattice by replacing Gd<sup>3+</sup> without destroying the tetragonal zircon type structure of GdVO<sub>4</sub>.

The morphology and size of obtained  $GdVO_4:30\%$  Eu NPs were investigated by scanning electron microscope (SEM). As shown in Figure 1A and B, it revealed that the resulting product consisted of uniform microerythrocyte-like NPs with diameter in a range of 200–300 nm. The dynamic light scattering (DLS) result also proved the conclusion (inset of Figure 1B). Further details of the sample structure were analyzed by transmission electron microscope (TEM) and high-resolution TEM (HRTEM). Figure 1C displays the TEM image of a typical GdVO<sub>4</sub>:30% Eu NPs with diameter of ca. 250 nm. In addition, the contrast between the relatively brighter centers and the dark surrounding area implies there was a depression in the center, supporting the claim of microarchitecture-like shape in SEM. As demonstrated by the corresponding HRTEM (inset of Figure 1C), it displayed well-defined lattice fringes with d spacing of 0.362 nm, which is in accordance with the lattice spacing in the (200) crystal planes of tetragonal GdVO<sub>4</sub>. Furthermore, the energy dispersive X-ray (EDX) result depicted in Figure S2 revealed that this sample was detected with significant quantities of  $Eu^{3+}$  in GdVO<sub>4</sub> NPs. To gain better insight into formation of the unique morphology of the GdVO<sub>4</sub> sample, we investigated the effect of sodium citrate serving as the chelating agent in the preparation process, as shown in the TEM images of the sample synthesized with different molar ratio of Na<sub>3</sub>VO<sub>4</sub>:Na<sub>3</sub>Cit. When decreasing the amount of Na<sub>3</sub>Cit, the Cit:Gd<sup>3+</sup> molar ratio was increased to 3:1, and the morphology of the sample was deduced to be a quasi-pancake hierarchical structure with a diameter below 200 nm (Figure 1D), with rough surface and evident boundaries of the obtained two-dimensional pancake-like patterns. Further decreasing the molar ratio to 2:1, imperfect microsphere



Figure I SEM images of GdVO<sub>4</sub>:30% Eu NPs. (A) Scale bar = 1  $\mu$ m and (B) scale bar = 100 nm. TEM images of GdVO<sub>4</sub>:30% Eu NPs with different molar of Cit<sup>3-</sup>/Gd<sup>3+</sup> (C) 4:1, (D) 3:1, (E) 2:1. Inset of b: DLS of the GdVO<sub>4</sub>:30% Eu NPs. Inset of c: HRTEM image of GdVO<sub>4</sub>:30% Eu NPs (scale bar = 5 nm).

structures were formed (Figure 1E) with rest of small NPs. Accordingly, citrate anions have a significant role in the formation of the  $GdVO_4$  microstructure.

In order to determine the optimal doping concentration, we varied the content of  $Eu^{3+}$  in the GdVO<sub>4</sub> samples. The dependence of the PL emission intensity on  $Eu^{3+}$  doping concentration (x) in Gd (1-x)VO<sub>4</sub>:xEu<sup>3+</sup> samples is displayed in Figure S3. It can be observed that the PL emission intensity of Eu<sup>3+</sup> doped GdVO<sub>4</sub> intensified with a corresponding increase in doping concentrations, and it leveled off when Eu<sup>3+</sup> concentration was over 20%. In view of concentration quenching effect, the optimum doping concentration for Eu<sup>3+</sup> was 30%. The PL behaviors of GdVO<sub>4</sub>:30% Eu NPs are illustrated in Figure 2A. The excitation spectrum consisted of a broad band centered at about 285 nm. The strong absorption in the UV range is ascribed to  $GdVO_4$  host  $(VO_4^{3-})$  excitation band due to V-O charge transfer inside the  $VO_4^{3-}$  groups, followed by transfer from the  $VO_4^{3-}$  groups to the Eu<sup>3+</sup> ions.<sup>30,31</sup> Sharp and well-resolved peaks ranging from 500-700 nm could be observed in the emission spectrum similar to the characteristic lanthanide luminescence, which originates from the  ${}^{5}D_{0} \rightarrow {}^{7}F_{J}$  (J = 1, 2, 3, 4) electronic transitions expected for Eu<sup>3+</sup>. Among them, the dominant red emission at 618 nm due to  ${}^{5}D_{0} \rightarrow {}^{7}F_{2}$  transition was higher than that associated with other transitions because of the absence of inversion center for the introduction of Eu<sup>3+</sup> ions which replace the normal sites of Gd<sup>3+</sup> ions in the tetragonal  $GdVO_4$  structure.<sup>32</sup> In addition, no emission from the  $VO_4^{3-}$  groups was observed, indicating the efficient energy transfer between VO4<sup>3-</sup> groups and Eu<sup>3+</sup> ions. A bright red fluorescence could be observed in the GdVO4:30% Eu NPs under a UV-light illumination. The PL lifetime of <sup>5</sup>D<sub>0</sub> (Eu<sup>3+</sup>) monitored at 618 nm for GdVO<sub>4</sub>:30% Eu NPs was determined to be 0.583 ms, as shown in Figure 2B. Such a long lifetime makes it possible for TR technique to improve the sensitivity of detection.

Furthermore, Fourier transform infrared (FTIR) spectrum of the GdVO<sub>4</sub>:30% Eu NPs sample was recorded in the wave number range of 400–4000 cm<sup>-1</sup> (Figure S4). The characteristic absorption peaks of GdVO<sub>4</sub> located at 807 and 450 cm<sup>-1</sup> can be ascribed to the stretching vibrations of the V-O (from the VO<sub>4</sub><sup>3-</sup> group) and Gd-O bond, respectively.<sup>33,34</sup> A broad absorption band at 3423 cm<sup>-1</sup> and a weak band at 1646 cm<sup>-1</sup> can be assigned to the O-H stretching and bending vibration of the water molecule physically absorbed on the sample.<sup>14</sup> The band at 1083 cm<sup>-1</sup> originates from the C-OH groups,<sup>35</sup> and those at 1560 and 1389 cm<sup>-1</sup> correspond to the asymmetric (v<sub>as</sub>) and symmetric (v<sub>s</sub>) stretching vibrations of the carboxylic group (-COO<sup>-</sup>), respectively.<sup>36</sup> These carboxylic groups were introduced due to some Cit<sup>3-</sup> molecules on the surface of the particles, which render the GdVO<sub>4</sub>:30% Eu NPs excellent water solubility and biocompatibility without any further surface functionalization, beneficial to their potential application as luminescent bio-labels. GdVO<sub>4</sub>:30% Eu NPs was –31.9 mV, further indicating the negatively charged carboxylic groups on the surface of GdVO<sub>4</sub>:30% Eu NPs. The zeta potential value varied to –9.63 mV after the conjugation with the Ab1. The antibody labeling concentration was determined with Bradford kit. Firstly, the calibration curve was established using a series of standards and the absorbance of Ab-conjugated GdVO<sub>4</sub>:30% Eu NPs with known concentration at 560 nm was



Figure 2 (A) The excitation spectrum (black line) and emission spectrum (red line) of GdVO<sub>4</sub>:30% Eu NPs. (B) The PL decay from  ${}^{5}D_{0}$  of the NPs emission monitoring at 618 nm.

measured. Then, according to the linear equation (A = 0.211C + 0.178, R<sup>2</sup> = 0.999) in <u>Figure S6</u>, the mass ratio of Ab/ NPs was 27 µg mg<sup>-1</sup>, and the labelling efficiency was calculated to be 40.5–54%.

## Analytical Performance of LFA Platform

On the basis of the prepared method, we assessed the working curve for the determination of the cTnI serum samples. The control solution without analytes showed no fluorescent signal on T line, which proved the results were valid (Figure 3A). With target-specific binding, the red PL of the T line gradually intensified with increase in concentrations of target cTnI under 310 nm ultraviolet illumination. Therefore, such a visible color change provides a qualitative or semiquantitative evaluation of cTnI level by the naked eve. It was satisfactory to find that this LFA strip was able to perform a visual readout even when the concentration of cTnI was as low as 0.025 ng mL<sup>-1</sup>. This visual detection limit was lower than 0.03 ng mL<sup>-1</sup> which is recommended as the positive threshold value for clinically applicable AMI diagnoses.<sup>4</sup> In order to further determine the sensitivity of the LFA system, standard solutions with known cTnI concentrations ranging from 0.02 to 30 ng mL<sup>-1</sup> were measured on the commercial fluorospectrometer. A good linear dependence was found between the fluorescent intensity ratio of T/C and cTnI concentration. The linear equation was fitted to be y = 0.0971x + 0.0371 with a correlation coefficient of 0.998 (Figure 3B). More importantly, LOD reached a value as low as  $17 \text{ pg mL}^{-1}$ . Compared with other reports related to cTnI determination (Table 1),<sup>13,37–45</sup> the prepared lanthanide-doped NPs-based LFA has the merits of relatively low LOD and wide linear range. As for colorimetry, it is hard to distinguish the intermediate color with the naked eye. However, just a portable hand-held fluorescent lamp can realize an objectively qualitative test on the prepared LFA even in an emergency situation or in areas/situations without specialized equipment. As is well known, in healthy individuals, the cTnI concentration is maintained at 0.1–0.3 ng mL<sup>-1</sup>, however the patients with suspected AMI are identified with the cut-off level of 6 ng mL<sup>-1</sup> at 1 hour.<sup>13</sup> In this regard, the developed platform is sensitive enough to be employed in clinical application. Also of note, this platform can complete all steps within 15 min for one sample, thus the proposed assay with combined high sensitivity and rapidity has competitive potential for on-site cTnI detection.

## Reproducibility, Specificity and Stability of LFA Platform

The reproducibility, specificity and stability of an immunoassay are important for analytical application. We measured the intra- and inter-assay variations to investigate the bias between different strips. Herein, intra-assay coefficient of variation (CV) and inter-assay CV were calculated as listed in Table 2, ranging from 3.67-6.86% and 4.92-7.92%, respectively. Both CVs were lower than 10%, validating the high precision of our LFA strip. In addition to reproducibility, another important issue that really matters in a newly developed platform is the specificity. To investigate this point, we tested the responses of LFA strip with cTnI samples containing a range of common interferents and cross-reactivity proteins in real serum samples. As depicted in Table 3, two samples with different cTnI concentrations all exhibited relative deviations (RD) within  $\pm 10\%$ ,





Table I	I Comparison	of Various Metho	ods for cTnl Dete	rmination
---------	--------------	------------------	-------------------	-----------

Materials	Detection Method	Linear Range	Detection Limit	Reference
15 nm AuNP-(ald) HRP-Ab and 40 nm AuNP-(ald)	LFA	10°–10 <sup>4</sup> and 10 <sup>1</sup> –	5.6 pg mL <sup><math>-1</math></sup> and 30	[37]
HRP-Ab		$10^4 \text{ pg mL}^{-1}$	pg mL <sup>-1</sup>	
Carbon nanofibers modified cellulose	LFA	0–250 ng mL <sup>-1</sup>	1.28–1.40 ng mL <sup>-1</sup>	[13]
ZrMOF@CdTe nanoparticles	LFA	0–1 mg L <sup>-1</sup>	250 ng mL $^{-1}$	[38]
Alexa Fluor dye	LFA	0–100 ng mL <sup>-1</sup>	19 pg mL <sup>-1</sup>	[39]
Polyvinyl alcohol	LFA	$10^{\circ}-10^{5} \text{ pg mL}^{-1}$	0.92 ng mL <sup>-1</sup>	[40]
Streptococcal protein G modified polystyrene	LFA	0.05–32 ng mL <sup>-1</sup>	32 pg mL <sup>-1</sup>	[41]
microspheres				
Raspberry-type Eu-SiNP	LFA	0–1.16 ng mL <sup>-1</sup>	97 pg mL <sup>-1</sup>	[42]
Peptide-functionalized gold nanoparticles	Colorimetry	0.1–1000 ng mL <sup>-1</sup>	0.2 ng mL <sup><math>-1</math></sup>	[43]
Gold nanorods (AuNRs) on heparin	Colorimetry	$0.5 - 15 \text{ ng mL}^{-1}$	0.4 ng mL <sup><math>-1</math></sup>	[44]
PDMS-AuNPs composite film	Colorimetry	$0.01-5 \text{ ng mL}^{-1}$	0.01 ng mL <sup>-1</sup>	[45]
GdVO₄:30% Eu NPs	LFA	0.02–30 ng mL <sup>-1</sup>	17 pg mL <sup>-1</sup>	This work

Table 2 The Reproducibility Results of the Developed LFA Strip

cTnI (ng mL <sup>−I</sup> )	Intra-Assay Precision (n=10)		Inter-Assay Precision (n=10)	
	Mean ± SD (ng mL <sup>-1</sup> )	CV (%)	Mean ± SD (ng mL <sup>-1</sup> )	CV (%)
0.62	0.55 ± 0.08	6.86	0.52 ± 0.11	6.92
2.62	2.53 ± 0.08	4.86	2.42 ± 0.11	4.92
8.45	8.19 ± 0.41	3.67	8.21 ± 0.67	5.26
20.29	21.36 ± 1.98	6.29	22.56 ± 2.63	7.92

Table 3 The Specificity Results of the Developed LFA Strip with Interferents on cTnl Detection

cTnI (ng mL <sup>−I</sup> )	cTnl (I.62 ng mL <sup>−I</sup> )		cTnl (2.93 ng mL <sup>−1</sup> )	
	Value	RD (%)	Value	RD (%)
Control	1.64	1.2	3.08	5.1
Bilirubin	1.72	6.2	3.15	7.5
Sodium azide	1.52	-6.2	2.84	-3.I
Cholesterol	1.69	4.3	2.98	1.7
Hemoglobin	1.53	-5.6	2.81	-4.1
Human serum albumin	1.25	2.5	2.54	3.5
Cardiac troponin T	1.68	3.6	3.56	2.8
Immunoglobulin G	1.85	-1.5	2.37	1.9

Note: RD = (Value-Standard value)/Standard value.

revealing that coexistence of these potential interferents did not show remarkable interference effects. Moreover, as shown in Figure 4, the stability experiment noted that GdVO<sub>4</sub>:30% Eu NPs retained their bioactivity when used in test strips. The fluorescence intensity decay rate was 1.2% after repeating 100 times, which is better in comparison to 2.4% in regard to polystyrene microsphere-based LFA strip. Therefore, the favorable good reproducibility, anti-interference ability and stability of our LFA strip were well fitted for the accurate determination of cTnI even in complex biological fluids.

## **Real Sample Analysis**

To demonstrate the reliability of direct quantitation of cTnI in complex media, we also performed in vitro cTnI detection in Beckman quality control serum samples. The fluorescent signal at different sites on the NC membrane



Figure 4 Fluorescence intensity decay rate of the LFA strip at different repeat times. Black dots: GdVO<sub>4</sub> NPs probe; red dots: Polystyrene microsphere.

was depicted as the scan profile in Figure 5A. The fluorescent signal ratio (T/C ratio) exhibited a good linear dependence on the cTnI concentration as represented in Figure 5B. The linear equation was  $y = 0.018x + 6.3*10^{-4}$ , with a correlation coefficient (R<sup>2</sup>) of 0.997. This result revealed that this LFA strip was capable of reliable cTnI determination in an unknown serum sample. To further evaluate the accuracy of the developed LFA strip, comparisons of the quantitative results between the LFA strip and the commercial chemiluminescence immunoassay kit were made. The results of 86 serum samples were chosen randomly to ensure consistency. Correlation of two methods was evaluated using Passing-Bablok linear regression (y = 0.92 x + 0.25), as shown in Figure 5C, a good agreement could be observed between our results with respect to those measured by commercial kit and a correlation coefficient (R<sup>2</sup>) of 0.964 was obtained. In addition, bias was investigated by the results from Bland-Altman plots (Figure 5D). The mean relative differences between two sets of analysis results were all less than 10%, validating good consistency between these two approaches.

#### Conclusion

In this work, we developed a practical GdVO<sub>4</sub>:30% Eu NPs-based LFA strip for rapid and sensitive detection of cTnI in real blood samples. The unique features of luminescent probe GdVO<sub>4</sub>:30% Eu NPs, under time-resolved detection mode without short-lived background interferences, allow for high sensitivity in cTnI detection, whose concentration can also be referred just by visual fluorescence readout even in an emergency situation without specialized equipment. The developed POC platform with ultralow detection limit is theoretically applicable to clinical application, and further practically demonstrated satisfactory results to meet the requirements for the determination of cTnI using clinical samples. Furthermore, a high correlation coefficient was achieved between our results with respect to those measured independently by using a commercial kit, demonstrating the accuracy and reliability of the proposed assay. By virtue of the remarkable advantages including visual readout detection, speediness, improved sensitivity, high specificity and good reproducibility, it could be concluded that this lanthanide-doped nanoprobes-based LFA strip would be a good candidate for next-generation POC diagnosis of biomarkers.



Figure 5 (A) The fluorescence signals of various concentrations of cTnI ranging from 0.1-30 ng mL<sup>-1</sup> spiked in Beckman cTnI quality control serum samples at different sites on the NC membrane. (B) Linear regression analysis of cTnI in Beckman quality control serum samples. (C) Correlation between the LFA kit based on GdVO<sub>4</sub>:30% Eu NPs and commercial kit for the determination of cTnI in 86 serum samples. (D) The difference between the developed LFA kit value and the commercial kit.

## **Supplementary Materials**

Additional figures for XRD, XPS, FTIR, ζ-potentials and calibration curve are provided.

#### **Acknowledgments**

The authors thank the Fujian Provincial University-Industry Cooperation Science & Technology Major Program (2019Y4006), Construction project of Fujian medical center of hematology (Min201704), Sponsored by National and Fujian Provincial Key Clinical Specialty Discipline Construction Program, P. R.C for financial support.

## Disclosure

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

## References

- 1. Mullins KE, Christenson RH. Optimal detection of acute myocardial injury and infarction with cardiac troponin: beyond the 99th Percentile, into the high-sensitivity era. *Curr Cardiol Rep.* 2020;22(9):1–10. doi:10.1007/s11886-020-01350-w
- Fathil MFM, Arshad MM, Gopinath SC, et al. Diagnostics on acute myocardial infarction: cardiac troponin biomarkers. *Biosens Bioelectron*. 2015;70:209–220. doi:10.1016/j.bios.2015.03.037
- 3. Han X, Li S, Peng Z, Othman AM, Leblanc RM. Recent development of cardiac troponin I detection. *Physiol Plant.* 2016;1(2):106–114. doi:10.1021/acssensors.5b00318
- Huang H, Zhu S, Wang W, et al. Diagnosis of acute myocardial infarction in patients with renal insufficiency using high-sensitivity troponin T. Clin Chem Lab Med. 2015;53(5):723–730. doi:10.1515/cclm-2014-0715
- 5. Staden SV, Ilie RM, Pogacean F, Pruneanu SM. Needle stochastic sensors for on-site fast recognition and quantification of biomarkers for gastric cancer in biological samples. *New J Chem.* 2020;44(46):20203–20211. doi:10.1039/D0NJ03847D
- 6. Regan B, O'Kennedy R, Collins D. Point-of-care compatibility of ultra-sensitive detection techniques for the cardiac biomarker troponin I-challenges and potential value. *Biosensors*. 2018;8(4):114. doi:10.3390/bios8040114

- Hassan SU, Tariq A, Noreen Z, Donia A, Zhang X. Capillary-driven flow microfluidics combined with smartphone detection: an emerging tool for point-of-care diagnostics. *Diagnostics*. 2020;10(8):509. doi:10.3390/diagnostics10080509
- 8. Wang C, Liu M, Wang Z, Li S, Deng Y, He N. Point-of-care diagnostics for infectious diseases: from methods to devices. *Nano Today*. 2021;37:101092. doi:10.1016/j.nantod.2021.101092
- 9. Yang J, Wang K, Xu H, Yan W, Jin Q, Cui D. Detection platforms for point-of-care testing based on colorimetric, luminescent and magnetic assays: a review. *Talanta*. 2019;202:96–110. doi:10.1016/j.talanta.2019.04.054
- 10. Jiang N, Ahmed R, Damayantharan M, Nal B, Yetisen AK. Lateral and vertical flow assays for point-of-care diagnostics. *Adv Healthcare Mater*. 2019;8(14):1900244. doi:10.1002/adhm.201900244
- 11. Guo J, Chen S, Tian S, Liu K, Ma X, Guo J. A sensitive and quantitative prognosis of C-reactive protein at picogram level using mesoporous silica encapsulated core-shell up-conversion nanoparticle based lateral flow strip assay. *Talanta*. 2021;230:122335. doi:10.1016/j.talanta.2021.122335
- 12. Bishop JD, Hsieh HV, Gasperino DJ, Weigl BH. Sensitivity enhancement in lateral flow assays: a systems perspective. Lab Chip. 2019;19 (15):2486-2499. doi:10.1039/C9LC00104B
- 13. Natarajan S, Jayaraj J, Prazeres DMF. A cellulose paper-based fluorescent lateral flow immunoassay for the quantitative detection of cardiac troponin I. *Biosensors*. 2021;11(2):49. doi:10.3390/bios11020049
- 14. Li Z, Wang Y, Wang J, Tang Z, Pounds JG, Lin Y. Rapid and sensitive detection of protein biomarker using a portable fluorescence biosensor based on quantum dots and a lateral flow test strip. *Anal Chem.* 2010;82(16):7008–7014. doi:10.1021/ac101405a
- 15. Liu Q, Cheng S, Chen R, et al. Near-infrared lanthanide-doped nanoparticles for a low interference lateral flow immunoassay test. ACS Appl Mater Inter. 2020;12(4):4358–4365. doi:10.1021/acsami.9b22449
- 16. Wilhelm S. Perspectives for upconverting nanoparticles. ACS Nano. 2017;11(11):10644. doi:10.1021/acsnano.7b07120
- 17. Liu QY, Cheng SM, Chen R, et al. Hybrids of upconversion nanoparticles and silver nanoclusters ensure superior bactericidal capability via combined sterilization. ACS Appl Mater Inter. 2020;12(46):51285-51292. doi:10.1021/acsami.0c15710
- 18. Chen ZH, Zhang ZG, Zhai XM, et al. Rapid and sensitive detection of anti-SARS-CoV-2 IgG, using lanthanide-doped nanoparticles-based lateral flow immunoassay. *Anal Chem.* 2020;92(10):7226–7231. doi:10.1021/acs.analchem.0c00784
- Zheng W, Tu D, Huang P, Zhou S, Chen Z, Chen X. Time-resolved luminescent biosensing based on inorganic lanthanide-doped nanoprobes. *Chem Commun.* 2015;51(20):4129–4143. doi:10.1039/C4CC10432C
- 20. da Rosa PPF, Kitagawa Y, Hasegawa Y. Luminescent lanthanide complex with seven-coordination geometry. *Coordin Chem Rev.* 2020;406:213153. doi:10.1016/j.ccr.2019.213153
- Liu Y, Tu D, Zhu H, Chen X. Lanthanide-doped luminescent nanoprobes: controlled synthesis, optical spectroscopy, and bioapplications. *Chem Soc Rev.* 2014;42(16):6924–6958. doi:10.1039/C3CS60060B
- 22. Escudero A, Carrillo-Carrión C, Zyuzin MV, Parak WJ. Luminescent rare-earth-based nanoparticles: a summarized overview of their synthesis, functionalization, and applications. *Top Curr Chem.* 2016;374(4):1–15. doi:10.1007/s41061-016-0049-8
- Mousseau F, Tarisse CF, Simon S, Gacoin T, Alexandrou A, Bouzigues CI. Luminescent lanthanide nanoparticle-based imaging enables ultra-sensitive, quantitative and multiplexed in vitro lateral flow immunoassays. *Nanoscale*. 2021;13(35):14814–14824. doi:10.1039/D1NR03358A
- Abdesselem M, Schoeffel M, Maurin I, et al. Correction to multifunctional rare-earth vanadate nanoparticles: luminescent labels, oxidant sensors, and MRI contrast agents. ACS Nano. 2015;9(4):4660. doi:10.1021/acsnano.5b01924
- 25. Kang X, Yang D, Dai Y, et al. Poly (acrylic acid) modified lanthanide-doped GdVO<sub>4</sub> hollow spheres for up-conversion cell imaging, MRI and pH-dependent drug release. *Nanoscale*. 2013;5(1):253–261. doi:10.1039/c2nr33130f
- Chen C, Li C, Li T, et al. Water-soluble, monodisperse, lanthanide-doped Y(Gd)VO<sub>4</sub> nanocrystals as promising multimodal bioprobe. *Eur J Inorg Chem.* 2015;2015(19):3108–3115. doi:10.1002/ejic.201500212
- Song Y, Shao B, Feng Y, et al. Emission enhancement and color tuning for GdVO<sub>4</sub>: Ln<sup>3+</sup> (Ln=Dy, Eu) by surface modification at single wavelength excitation. *Inorg Chem.* 2017;56(1):282–291. doi:10.1021/acs.inorgchem.6b02125
- Kaowphong S, Chumha N, Nimmanpipug P, Kittiwachana S. Nanosized GdVO<sub>4</sub> powders synthesized by sol-gel method using different carboxylic acids. *Rare Met.* 2018;37(7):561–567. doi:10.1007/s12598-015-0681-4
- 29. Jin D, Yang H, Ding G, Yu X, Wang L, Zheng Y. Hydrothermal synthesis and photoluminescence behavior of Eu-doped GdVO4. *Inorg Mater*. 2008;44(10):1121–1124. doi:10.1134/S002016850810018X
- 30. Li X, Yu M, Hou Z, et al. One-dimensional GdVO<sub>4</sub>:Ln<sup>3+</sup> (Ln=Eu, Dy, Sm) nanofibers: electrospinning preparation and luminescence properties. J Solid State Chem. 2011;184(1):141–148. doi:10.1016/j.jssc.2010.11.019
- 31. Szczeszak A, Grzyb T, Śniadecki Z, et al. Structural, spectroscopic, and magnetic properties of Eu<sup>3+</sup>-doped GdVO<sub>4</sub> nanocrystals synthesized by a hydrothermal method. *Inorg Chem.* 2014;53(23):12243–12252. doi:10.1021/ic500354t
- 32. Liu Y, Liu G, Dong X, Wang J, Yu W. Tunable photoluminescence and magnetic properties of Dy<sup>3+</sup> and Eu<sup>3+</sup> doped GdVO<sub>4</sub> multifunctional phosphors. *Phys Chem Chem Phys.* 2015;17(40):26638–26644. doi:10.1039/c5cp04373e
- 33. Liu G, Hong G, Wang J, Dong X. Deposition of GdVO<sub>4</sub>:Eu<sup>3+</sup> nanoparticles on silica nanospheres by a simple sol-gel method. *Nanotechnology*. 2006;17(13):3134. doi:10.1088/0957-4484/17/13/009
- 34. Xu Z, Feng B, Gao Y, et al. Uniform and well-dispersed GdVO<sub>4</sub> hierarchical architectures: hydrothermal synthesis, morphology evolution, and luminescence properties. *CrystEngComm*. 2012;14(17):5530–5538. doi:10.1039/c2ce25521a
- 35. Gabrienko AA, Ewing AV, Chibiryaev AM, Agafontsev AM, Dubkov KA, Kazarian SG. New insights into the mechanism of interaction between CO<sub>2</sub> and polymers from thermodynamic parameters obtained by in situ ATR-FTIR spectroscopy. *Phys Chem Chem Phys.* 2016;18(9):6465–6475. doi:10.1039/c5cp06431g
- 36. Yang L, Li L, Zhao M, Li G. Size-induced variations in bulk/surface structures and their impact on photoluminescence properties of GdVO<sub>4</sub>:Eu<sup>3</sup> <sup>+</sup>nanoparticles. *Phys Chem Chem Phys.* 2012;14(28):9956–9965. doi:10.1039/c2cp41136a
- Han GR, Kim MG. Highly sensitive chemiluminescence-based lateral flow immunoassay for cardiac troponin I detection in human serum. Sensors. 2020;20(9):2593. doi:10.3390/s20092593
- 38. Zou J, Liu X, Ren X, et al. Rapid and simultaneous detection of heart-type fatty acid binding protein and cardiac troponin using a lateral flow assay based on metal organic framework@CdTe nanoparticles. *Nanoscale*. 2021;13(16):7844–7850. doi:10.1039/D1NR00702E
- Natarajan S, Su F, Jayaraj J, Shah MII, Huang Y. A paper microfluidics-based fluorescent lateral flow immunoassay for point-of-care diagnostics of non-communicable diseases. *Analyst.* 2019;144(21):6291–6303. doi:10.1039/C9AN01382B

- 40. Han GR, Koo HJ, Ki H, Kim MG. Paper/soluble polymer hybrid-based lateral flow biosensing platform for high-performance point-of-care testing. *ACS Appl Mater Inter*. 2020;12(31):34564–34575. doi:10.1021/acsami.0c07893
- 41. Lou D, Fan L, Cui Y, Zhu Y, Gu N, Zhang Y. Fluorescent nanoprobes with oriented modified antibodies to improve lateral flow immunoassay of cardiac troponin I. *Anal Chem.* 2018;90(11):6502–6508. doi:10.1021/acs.analchem.7b05410
- 42. Lee KW, Kim KR, Chun HJ, et al. Time-resolved fluorescence resonance energy transfer-based lateral flow immunoassay using a raspberry-type europium particle and a single membrane for the detection of cardiac troponin I. *Biosens Bioelectron*. 2020;163:112284. doi:10.1016/j. bios.2020.112284
- 43. Liu X, Wang Y, Chen P, et al. Peptide functionalized gold nanoparticles with optimized particle size and concentration for colorimetric assay development: detection of cardiac troponin I. Acs Sensors. 2016;1(12):1416–1422. doi:10.1021/acssensors.6b00493
- 44. Raj V, Alex S. Non-enzymatic colorimetric sensor for cardiac Troponin I (cTnI) based on self-assembly of gold nanorods on heparin. *Gold Bull*. 2021;54(1):1–7. doi:10.1007/s13404-020-00287-w
- 45. Wu WY, Bian ZP, Wang W, Zhu JJ. PDMS gold nanoparticle composite film-based silver enhanced colorimetric detection of cardiac troponin I. Sens Actuators B Chem. 2010;147(1):298–303. doi:10.1016/j.snb.2010.03.027

International Journal of Nanomedicine

#### **Dove**press

Publish your work in this journal

The International Journal of Nanomedicine is an international, peer-reviewed journal focusing on the application of nanotechnology in diagnostics, therapeutics, and drug delivery systems throughout the biomedical field. This journal is indexed on PubMed Central, MedLine, CAS, SciSearch<sup>®</sup>, Current Contents<sup>®</sup>/Clinical Medicine, Journal Citation Reports/Science Edition, EMBase, Scopus and the Elsevier Bibliographic databases. The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit http:// www.dovepress.com/testimonials.php to read real quotes from published authors.

Submit your manuscript here: https://www.dovepress.com/international-journal-of-nanomedicine-journal