Comparison of CDE data in phacoemulsification between an open hospital-based ambulatory surgical center and a free-standing ambulatory surgical center

Ming Chen¹
Mindy Chen²

¹University of Hawaii, Honolulu, HI, USA; ²University of California, Irvine, CA, USA

Abstract: Mean CDE (cumulative dissipated energy) values were compared for an open hospital-based surgical center and a free-standing surgical center. The same model of phacoemulsifier (Alcon Infiniti Ozil) was used. Mean CDE values showed that surgeons (individual private practice) at the free-standing surgical center were more efficient than surgeons (individual private practice) at the open hospital-based surgical center (mean CDE at the hospital-based surgical center 18.96 seconds [SD = 12.51]; mean CDE at the free-standing surgical center 13.2 seconds [SD = 9.5]). CDE can be used to monitor the efficiency of a cataract surgeon and surgical center in phacoemulsification. The CDE value may be used by institutions as one of the indicators for quality control and audit in phacoemulsification.

Keywords: CDE (cumulative dissipated energy), open hospital-based ambulatory surgical center, free-standing surgical center, phacoemulsification

Introduction

CDE (cumulative dissipated energy) is a built-in device of the Alcon Infiniti phacoemulsification unit (Alcon Labs, Hünenberg, Switzerland). It shows the time taken in seconds to complete a case of phacoemulsification, which reflects how much energy will be expended. Less phacoemulsification time translates to less energy used in the eye, and is considered better for corneal endothelium.¹⁻⁴ The purpose of this study was to introduce the CDE as a monitor or measure of surgical efficiency in the hope of improving surgical outcomes. CDE data are not currently being widely utilized for this purpose. A higher CDE reading is equated with longer surgical and recovery time, because more energy is dissipated in the eye, resulting in more damage to ocular tissue. Successful phacoemulsification of the cataract with less CDE may correlate with more efficient surgery and better overall outcomes. A previous study used CDE to compare torsional mode with conventional ultrasound mode phacoemulsification.⁵ Another study utilized the CDE to compare different techniques and settings by different surgeons in phacoemulsification.⁶

Objective

This study set out to compare mean CDE between surgeons in an open hospital-based surgical center and a free-standing surgical center, using the same model of phacoemulsifier (Alcon Infiniti Ozil). It also intended to establish the mean CDE

Correspondence: Ming Chen
55 S. Kukui St C-109, Honolulu,
Hawaii, HI 96813, USA.
Tel: +1 808 531 8874
Fax: +1 808 523 0466
Email: mingchen@hawaii.rr.com

value in the two institutions as a reference for future training and quality control in cataract surgery.

Method

Ninety cases were retrospectively selected from the records of each institution's operating room records in one randomly selected month. Because all 180 cases were selected from one randomly selected month of the year and from similar ambulatory settings of the two nearby ambulatory surgical centers with similar descriptive data, the variety of cataract cases (soft vs hard) was expected to be similar. CDE value, taken from the phacoemulsifier, was recorded by the nurse after phacoemulsification. The CDE values of 90 cases from each institution were recorded in an Excel spreadsheet together with age and gender. There is no necessity to record the patient's identity, which avoids privacy intrusion. The data were then analyzed by SPSS. The ratio of males and females was equal at the hospital and 45.2% versus 54.8% at the surgical center. The age distribution was from 28 to 94 years for the hospital versus 20 to 99 years for the surgical center.

This study was approved by the IRB (Institutional Review Board) of the University of Hawaii for ethical issues.

Result

The mean CDE at the hospital-based center was 18.96 seconds (SD = 12.51); for the surgical center it was 13.2 seconds (SD = 9.5) (Figure 1).

Conclusion

CDE values showed that surgeons at the free-standing ambulatory surgical center were more efficient in phacoemulsification than surgeons at the open hospital-based ambulatory surgical center. The free-standing cataract surgical center may be more efficient in cataract surgery than the open hospital-based ambulatory surgical center.

Discussion

This study intended to use CDE as an objective guide to evaluate the efficiency of phacoemulsification. The open hospital-based ambulatory surgical center has more surgeons (individual private practice) on staff, and each surgeon has a different volume of cataract surgeries. The free-standing cataract surgical center only has four surgeons on staff (individual private practice), and each surgeon has a uniformly large volume of cataract surgeries. Furthermore, the majority of the free-standing surgical center's cases are cataract surgeries, whereas the hospital-based ambulatory surgical center deals with a variety of different eye cases. Even though the hospital-based ambulatory surgical center is hospital based, it is still an ambulatory center. It is designed for ambulatory care to follow the ambulatory surgery guidelines, as with the free-standing surgical center. Because the cataract surgery was not performed in the main operation room of the hospital but rather in the ambulatory center, cataract cases selected should not be more difficult than those in the

Figure 1 Mean CDE in hospital-based ambulatory and free-standing centers.

free-standing surgical center. Because all 180 cases were selected randomly from one month of the year and from similar ambulatory settings of the two nearby ambulatory surgical centers with similar descriptive data, the variety of cataract cases (soft vs hard) was expected to be similar. This study showed that the high-volume cataract surgeons at the free-standing surgical center had a low mean CDE because of their better efficiency in phacoemulsification. CDE can be used as a monitor for a cataract surgeons' efficiency in phacoemulsification. A mean CDE value was established for the two institutions in this study. The CDE value may be used by institutions as an indicator of quality control and audit in phacoemulsification.

Disclosure

The authors disclose no conflicts of interest.

References

1. Pirazzoli B, D'Eliseo D, Ziosi M, Acciarri R. Effects of phacoemulsification time on the corneal endothelium using phacofracture and phaco chop techniques. *J Cataract Refract Surg*. 1996;22(7):967–969.
2. Kreisler KR, Mortenson SW, Mamalis N. Endothelial cell loss following “modern” phacoemulsification by a senior resident. *Ophthalmic Surg*. 1992;23(3):158–160.
3. Colvard DM, Kratz RP, Mazzocco TR, Davidson B. Endothelial cell loss following phacoemulsification in the pupillary plane. *Am Intra-Ocular Implant Soc J*. 1981;7(4):334–336.
4. Hayashi K, Hayashi H, Hakao F, and Kayashi F. Risk factors for corneal endothelial injury during phacoemulsification. *J Cataract Refract Surg*. 1996;22(8):1079–1084.
5. Liu Y, Zeng M, Liu X, et al. Torsional mode versus conventional ultrasound mode phacoemulsification. *J Cataract Refract Surg*. 2007;33(2):287–292.
6. Chen M, Sweeney H, Luke B, Chen M, et al. A retrospective randomized study to compare the energy delivered using CDE with different techniques and OZil® settings by different surgeons in phacoemulsification, *Clinical Ophthalmology*, 2009;3:401–403.

Clinical Ophthalmology

Publish your work in this journal

Clinical Ophthalmology is an international, peer-reviewed journal covering all subspecialties within ophthalmology. Key topics include: Optometry; Visual science; Pharmacology and drug therapy in eye diseases; Basic Sciences; Primary and Secondary eye care; Patient Safety and Quality of Care Improvements. This journal is indexed on

Submit your manuscript here: <http://www.dovepress.com/clinical-ophthalmology-journal>

Dovepress

PubMed Central and CAS, and is the official journal of The Society of Clinical Ophthalmology (SCO). The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit <http://www.dovepress.com/testimonials.php> to read real quotes from published authors.